

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

info@seamannuclear.com, Tel 414-762-5100 Fax 414-762-5108

Leak Test Certificate

April 10, 2012

A leak test has been performed on meter, serial number **8172**, a model R-50, containing the radionuclide Am 241.

Owned By: Thermoscan Inspections Inc.
P.O. Box 705
Carmel, IN 46082


Date Sample Collected: 4/6/12
Collected By: Pam
Date Sample Analyzed: 4/10/12
Analyzed By: T. Lindemann

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 10/6/12



Scott C. Seaman
President & Radiation Safety Officer, Seaman Nuclear Corporation

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
info@seamannuclear.com, Tel 414-762-5100 Fax 414-762-5100

Leak Test Certificate

April 10, 2012

A leak test has been performed on meter, serial number **8152**, a model R-50, containing the radionuclide Am 241.

Owned By: Thermoscan Inspections Inc.
P.O. Box 705
Carmel, IN 46082


Date Sample Collected: 4/6/12
Collected By: Pam
Date Sample Analyzed: 4/10/12
Analyzed By: 8172T. Lindemann

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 10/6/12



Scott C. Seaman
President & Radiation Safety Officer, Seaman Nuclear Corporation

**Fax Transmittal Form**

US NRC
To: Toye Simmons
630-515-1078

From: Lynne Easton

Date sent: 6-29-12
Number of pages including cover page: 3

Message:

Ms. Simmons,

Here are the last two leak
test certificates. Please let
me know if you need anything
else.

Thanks!
Lynne

Thermo-Scan Inspections

P.O. Box 705
Carmel, IN 46082
Phone: 317-846-4655
Toll Free: 1-800-481-5748
Fax: 317-846-4672