NRC FORM 591M PART 1
(05-2012)
10 CFR 2.201

U.S. NUCLEAR REGULATORY COMMISSION

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:	2. NRC/REGIONAL OFFICE
Scheurer Hospital 170 N. Caseville Road Pigeon, MI 48755 REPORT NUMBER(S) 2012-001	Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352

3. DOCKET NUMBER(S)	4. LICENSE NUMBER(S)	5. DATE(S) OF INSPECTION
030-35444	21-32250-01	June 26, 2012

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

X 1. Based on the inspection findings, no violations were identified.

2. Previous violation(s) closed.

3. The violations(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura, Health Physicist	D Piskura	6/26/12
BRANCH CHIEF	Hironori Peterson, Acting Chief, MIB		7/9/12

NRC FORM 591M PART 3
(05-2012)
10 CFR 2.201

U.S. NUCLEAR REGULATORY COMMISSION

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:	2. NRC/REGIONAL OFFICE
Scheurer Hospital 170 N. Caseville Road Pigeon, MI 48755 REPORT NUMBER(S) 2012-001	Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352

3. DOCKET NUMBER(S)	4. LICENSE NUMBER(S)	5. DATE(S) OF INSPECTION
030-35444	21-32250-01	June 26, 2012

6. INSPECTION PROCEDURES USED	7. INSPECTION FOCUS AREAS
87130, 87131	03.01 - 03.07

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)	2. PRIORITY	3. LICENSEE CONTACT	4. TELEPHONE NUMBER
02120	3	Mark Guilfoyle, D.O., RSO	(989) 453-3223

☑ Main Office Inspection Next Inspection Date: June 2015

☐ Field Office Inspection

☐ Temporary Job Site Inspection

PROGRAM SCOPE

This licensee was a small community hospital and authorized to use licensed material permitted by Sections 35.100, 35.200, and 35.300. The nuclear medicine department was staffed with one full-time technologist who performed approximately 80-100 diagnostic procedures monthly. These included a full spectrum of studies. The licensee received unit doses from a licensed radiopharmacy. The licensee administered therapeutic radiopharacueticals on one or two occasions annually; these cases were limited to hyperthyroid treatments. The licensee obtained its radioiodine in capsule form. The licensee retained the services of a consultant physicist who audited the radiation safety program quarterly (last 4/5/2012).

This inspection consisted of interviews with select licensee personnel; a review of select records; a tour of the nuclear medicine department; and independent measurements. The inspector observed the administration of one procedure for a gall bladder study. The inspection included observations of dose calibrator QA checks, security of byproduct material, use of personnel monitoring, and surveys.