

Hammond, Michelle

From: Hammond, Michelle
Sent: Wednesday, July 11, 2012 9:20 AM
To: 'jpasqua@hrrsa.gov'
Cc: 'acorey@hrrsa.gov'
Subject: Re: REQUEST FOR ADDITIONAL INFORMATION
Attachments: sr1556v7.pdf; image001.png

License No.: 17-14996-01
Docket No.: 030-08380
Control No.: 576956

Dear Mr. Pasqua:

This is in reference to your license application for renewal dated February 10, 2012 regarding Nuclear Regulatory Commission License No. 17-14996-01. In order to continue our review, we need the following additional information:

- 1) Please refer to the attached document and complete the checklist in Appendix C (pgs 93-102). Please Sign, date, scan as pdf. and include in your response.
- 2) Please provide the manufacturers and model number of the Cs-137 sealed source requested in Item 5 of your application (Item 5 of checklist re: sealed sources).
- 3) Please provide the additional information regarding facilities and equipment in Item 9 of the application (see Item 9 checklist and Section 8.9 –Response from applicant). Specifically, the safety and detection equipment that will be used in Room 3520C where P-32 and P-33 will be used, as well as any engineering controls that will be used when working with I-131.
- 4) Please provide information of how the radioactive material will be secured in labs accessible by students, faculty, and ancillary personnel (in accordance with 10 CFR 20.1801 and 20.1802). A review of your inspection history shows a violation regarding the security of RAM and we would like to ensure that corrective actions have been implemented.

We will continue our review upon receipt of this information. Please reply to my attention at the Region IV Office and refer to Mail Control No. 576956. Please respond to this e-mail by **July 25, 2012**. You may reply via e-mail as long as the response is attached, signed, dated, and in a PDF format or by fax to 817-200-1188.

Please don't hesitate to contact me if you have any questions.

Thank You.

Best Regards,

*Michelle M. Hammond, M.Sc.
Health Physicist
Region IV-Division of Nuclear Materials Safety, Branch B
office) 817-200-1127
fax) 817-200-1188*

