

Hammond, Michelle

From: Hammond, Michelle
Sent: Monday, July 09, 2012 3:19 PM
To: 'mkrohn@amrad.com'
Subject: Re: REQUEST FOR ADDITIONAL INFORMATION
Attachments: image001.png; Leak Test chklist Vol 18.pdf

License No.: 17-29441-01
Docket No.: 030-38546
Control No.: 577458

Dear Mr. Krohn:

This is in reference to your license application dated April 30, 2012 regarding Nuclear Regulatory Commission License No. 17-29441-01. In order to continue our review, we need the following additional information:

- 1) Please confirm that you do not wish to possess the clients material. Confirm that the clients radioactive material will remain under the clients authorized possession, and that you will use the radioactive material to perform D&D activities under your proposed license.
- 2) Please commit to the following: *Pursuant to 10 CFR 30.35(g), we shall maintain drawings and records important to decommissioning and transfer these records to a new licensee before licensed activities are transferred, or assign the records to the appropriate NRC Regional Office before the license is terminated.*
- 3) Please provide procedures regarding the following: criteria and use of remote handling tools, sealed source rupture and/or if a sealed source is stuck in a housing. If you will not be performing such services, Please state that such work will be performed by a licensed equipment manufacturer.
- 4) Please provide a procedure for how measurements of radioactive material concentrations in water that is released to the environment or sanitary sewer will meet the requirements in 10 CFR Part 20.2003.
- 5) Please refer to the attached checklist regarding Leak Tests, check the appropriate boxes, and provide the checklist in your response.

We will continue our review upon receipt of this information. Please reply to my attention at the Region IV Office and refer to Mail Control No. 577458. Please respond to this e-mail by **July 13, 2012**. You may reply via e-mail as long as the response is attached, signed, dated, and in a PDF format or by fax to 817-200-1188.

Please don't hesitate to contact me if you have any questions.

Thank You.

Best Regards,

*Michelle M. Hammond, M.Sc.
Health Physicist
Region IV-Division of Nuclear Materials Safety, Branch B
office) 817-200-1127
fax) 817-200-1188*