



July 12, 2012

Dennis P. O'Dowd
Health Physicist
UNITES STATES NUCLEAR REGULATORY COMMISSION
Region III, Materials Licensing Section
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

Re: Clarification for control number: 577513

Mr. O'Dowd,

Please replace the document dated 5/9/2006 with the faxed document date 03/29/2012. The original document was sent in error. The faxed document is the individual results from the wipe tests conducted on 03/29/2012. The data corresponds to the map that was included in the original request.

Please note the sealed source inventory and leak tests were included as informational documentation only. There were no sealed sources stored in the treadmill area.

Thank you for your cooperation with this matter. If you have any question please contact me at (734) 662-3197.

Sincerely,

A handwritten signature in black ink, reading "Michelle L. Kritzman". The signature is fluid and cursive, with the first name "Michelle" being more prominent.

Michelle L. Kritzman, M.S.
Radiological Physicist
ABR Certified, Medical Nuclear Physics



Community Health Center of Branch County**License No. 21-16642-01****Date:** 3/29/2012**Performed By:** Linda Brown, RSO**Survey Meter:** Ludlum 14C**Check Value:** 6.5 mR/hr

		cpm	net dpm	mR/h
1	Desk	653	0	0.02
2	Floor	703	100	0.02
3	Floor	703	100	0.02
4	Floor	703	100	0.02
5	Floor	703	100	0.02
6	Floor	703	100	0.02
7	Floor	703	100	0.02
8	Table	668	13	0.02
	Background	663		0.02

ACTION LEVELS: 2000 dpm

Conversion factor 2.5 dpm/cpm


COMMENTS: No evidence of removable contamination.


**TELECOPY TRANSMITTAL**DATE: 07/13/12TO: Dennis O'DowdFax Number: 630-515-1078FROM: Michelle KritzmanRE: Additional Information for control #577513


COMMENTS:

This transmittal consists of 3 pages (including transmittal cover sheet). Please contact me if you have any questions or problems with this transmittal.

214 E. Huron Street
Ann Arbor, MI 48104
(734) 662-9224 Fax
(734) 662-3197

 70 E. 91st Street, Suite 106
Indianapolis, IN 46240
(317) 581-1931 Fax
(317) 581-1911

 4806 Mile High Drive
Salt Lake City, UT 84124
(801) 272-2952 Office & Fax

 N7375 Crystal Ridge Drive
Beaver Dam, WI 53916
(920) 885-9872 Fax
(920) 885-9870