

July 16, 2012

EN 47938 (closed)  
NMED 120313 (closed)

Mr. Wallace P. Millet  
Radiation Safety Officer  
Hemlock Semiconductor  
12334 Geddes Rd,  
Hemlock, MI 48626

SUBJECT: ERRATA FOR RESPONSE TO STUCK SHUTTER ON THE OHMART VEGA  
GAUGE EVENT DATED MAY 18, 2012.

Dear Mr. Millet:

This is an errata to correct the June 20, 2012, U. S. Nuclear Regulatory Commission (NRC) acknowledgement letter for your 30 day response to the stuck shutter event dated May 18, 2012. A paragraph in our acknowledgement letter inadvertently included a reference to a lost device.

Enclosed is the revised letter that correctly references your response to a stuck shutter event. No findings were identified. We regret any inconvenience this may have caused.

In accordance with Title 10 of the Code of Federal Regulations (CFR) 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Should you have any questions concerning this errata, we will be pleased to discuss them with you.

Sincerely,

**/RA/**

Tamara E. Bloomer, Chief  
Materials Inspection Branch  
Division of Nuclear Materials Safety

Docket No. 030-37611  
License No. 21-32682-01

Enclosure:  
As stated

cc w/encl: State of Michigan

EN 47938 (closed)  
NMED 120313 (closed)

Mr. Wallace P. Millet  
Radiation Safety Officer  
Hemlock Semiconductor  
12334 Geddes Rd,  
Hemlock, MI 48626

SUBJECT: ERRATA FOR RESPONSE TO STUCK SHUTTER ON THE OHMART VEGA  
GAUGE EVENT DATED MAY 18, 2012.

Dear Mr. Millet:

This is an errata to correct the June 20, 2012, U. S. Nuclear Regulatory Commission (NRC) acknowledgement letter for your 30 day response to the stuck shutter event dated May 18, 2012. A paragraph in our acknowledgement letter inadvertently included a reference to a lost device.

Enclosed is the revised letter that correctly references your response to a stuck shutter event. No findings were identified. We regret any inconvenience this may have caused.

In accordance with Title 10 of the Code of Federal Regulations (CFR) 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Should you have any questions concerning this errata, we will be pleased to discuss them with you.

Sincerely,

/RA/

Tamara E. Bloomer, Chief  
Materials Inspection Branch  
Division of Nuclear Materials Safety

Docket No. 030-37611  
License No. 21-32682-01

Enclosure:  
As stated

cc w/encl: State of Michigan

\*See previous concurrence

**ADAMS Accession Number: ML12200A348**

DOCUMENT NAME: G:\DNMS\Work in progress\LTR- Hemlock Semiconductor Cerrataltr rev 1.docx

☒ Publicly Available ☐ Non-Publicly Available ☐ Sensitive ☒ Non-Sensitive

To receive a copy of this document, indicate in the concurrence box "C" = Copy without attach/encl "E" = Copy with attach/encl "N" = No copy

OFFICE	RIII		RIII		RIII		RIII	
NAME	ELKulzer:jc TEB for		TEBloomer TEB for					
DATE	07/16/2012		07/16/2012					

OFFICIAL RECORD COPY

June 20, 2012

EN-47938 (closed)  
NMED No. 120313 (closed)

Mr. Wallace P. Millet  
Radiation Safety Officer  
Hemlock Semiconductor  
12334 Geddes Rd,  
Hemlock, MI 48626

SUBJECT: RESPONSE TO STUCK SHUTTER ON THE OHMART VEGA GAUGE EVENT  
DATED MAY 18, 2012

Dear Mr. Millet:

This acknowledges receipt of your letter to the U.S. Nuclear Regulatory Commission (NRC) dated May 18, 2012, reporting the stuck shutter on an Ohmart Vega gauge.

We have reviewed your letter, including your actions taken to secure and leak test the gauge and work with the NRC in finding the original licensee. Your action to have the Manufactures' Field Engineer fix the shutter appears to be adequate, and the NRC has no further questions at this time.

In accordance with Title 10 of the Code of Federal Regulations 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Assess and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

**/RA/**

Hironori Peterson, Acting Chief  
Materials Inspection Branch  
Division of Nuclear Materials Safety

Docket No. 030-37611  
License No. 21-32682-01

Enclosure