

**From:** [Weidner, Tara](#)  
**To:** ["babinskim@nanticoke.org"](mailto:babinskim@nanticoke.org)  
**Subject:** NRC - Request for information  
**Date:** Thursday, August 02, 2012 2:19:00 PM

---

License No. 07-17618-01  
Docket No. 030-13060  
Control No. 577948

**SUBJECT:** NANTICOKE MEMORIAL HOSPITAL, REQUEST FOR ADDITIONAL INFORMATION  
CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL NO. 577948

\*\*\*\*\* PLEASE RESPOND TO CONFIRM RECEIPT OF THIS REQUEST \*\*\*\*\*

Dear Ms. Babinski:

This is in reference to your letter dated July 20, 2012 requesting to amend Nuclear Regulatory Commission License No. 07-17618-01. In order to continue our review, we need the following additional information:

1. With regards to the PET/CT coach, confirm:

- a. the address where the coach will be permanently parked,
- b. that the coach will not be parked in a public right-of-way,
- c. the coach will not be moved without prior approval of the NRC,
- d. the location of use is not a location of use on any other active NRC license, and
- e. the location of use is property owned by Nanticoke Memorial Hospital (NMH). If the space is leased from another party, provide documentation of a clear contractual agreement concerning access to your location of use for the purpose of decontamination or removal of licensed material from the location of use in the event of disharmony between you and the owner entity. This documentation should consist of signed certification from both parties.

2. Please confirm that the PET/CT coach is owned by NMH and will be staffed by NMH personnel.

3. Confirm that the Radiation Safety Program established at NMH will be implemented at the new location of use.

4. Describe your methods for securing byproduct material (sealed sources, waste, doses, etc.) from unauthorized access within the PET/CT coach.

5. Please confirm if there are any NARM sealed sources or byproduct material that would not fall under 35.200 or 35.65.

6. Please identify any additional authorized users involved with your PET/CT program and

provide their training and experience as described in 10 CFR 35.290. NRC Form 313A (AUD) may be used to document this information.

We will continue our review upon receipt of this information. Please fax the signed response (referencing Mail Control No. 577948) to Penny Lanzisera at 610-337-5269. If you have any technical questions regarding this deficiency letter, please call Ms. Lanzisera at (610) 337-5169.

In order to complete the amendment by your requested date of August 10, 2012, we will need your reply by Tuesday, August 7.

*Tara L. Weidner*  
*Health Physicist*  
*US Nuclear Regulatory Commission*