

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Evans Analytical Group, LLC
EAG Life Sciences
2662 Metro Blvd.
Maryland Heights, MO 63042

REPORT NUMBER(S) 2012-001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-38316

4. LICENSE NUMBER(S)

24-32801-01

5. DATE(S) OF INSPECTION

July 31, thru Aug. 16, 2012

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura, Health Physicist		8/16/2012
BRANCH CHIEF	Tamara E. Bloomer, Chief, MIB		8/17/12

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Evans Analytical Group, LLC
EAG Life Sciences
2662 Metro Blvd.
Maryland Heights, MO 63042

REPORT NUMBER(S) 2012-001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-38316

4. LICENSE NUMBER(S)

24-32801-01

5. DATE(S) OF INSPECTION

July 31, thru Aug. 16, 2012

6. INSPECTION PROCEDURES USED

87126

7. INSPECTION FOCUS AREAS

03.01-03.08

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)

03620

2. PRIORITY

5

3. LICENSEE CONTACT

Michael Rogers, RSO

4. TELEPHONE NUMBER

(800) 659-6059

☒ Main Office Inspection

Next Inspection Date: August 2013

☐ Field Office Inspection

☐ Temporary Job Site Inspection

PROGRAM SCOPE

This licensee was a small biomedical research firm authorized to use H-3 and C-14 for in vitro laboratory R&D. At the time of this inspection, the licensee had neither possessed nor initiated use of radioactive material authorized by this license. The licensee possessed two generally licensed gas chromatographs which were used to analyze samples. The licensee anticipated use of C-14 and/or H-3 in the future (within 6-12 months), once certain research protocols were coordinated with its California lab.

Since no licensed activities were initiated since the previous inspection, the next inspection date was set for one year as required by MC 2800. This inspection included in-office review to discuss the licensee's anticipated use of RAM and the need to file an amendment to its license to reflect the current status. On August 15, 2012, the licensee filed an amendment request "for storage only."