



Department of State Health Services

RADIOACTIVE MATERIAL LICENSE

Pursuant to the Texas Radiation Control Act and Texas Department of State Health Services (Agency) regulations on radiation, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess and transfer radioactive material listed below; and to use such radioactive material for the purpose(s) and at the place(s) designated below. This license is subject to all applicable rules, regulations and orders of the Agency now or hereafter in effect and to any conditions specified below.

LICENSEE			This license is issued in response to the letter	
1. Name TRUGLO INC			Dated: January 29, 2013	
ATTN LORRAINE L HELLINGHAUSEN			Signed by: Lorraine L. Hellinghausen	
2. Address 525 INTERNATIONAL PKWY			3. License Number L05519	
RICHARDSON TX 75081			Amendment Number 09	
PREVIOUS AMENDMENTS ARE VOID				
			4. Expiration Date March 31, 2022	
RADIOACTIVE MATERIAL AUTHORIZED				
5. Radioisotope A. H-3	6. Form of Material A. Sealed source (MB-Microtec Models 400/1, 400/2 and 400/3)	7. Maximum Activity A. 60,000 sources not to exceed 30 millicuries each	8. Authorized Use A. Manufacture and storage of archery pins and gun sights containing self-luminous light sources pending distribution pursuant to U.S. Nuclear Regulatory Commission License No. 42-23889-01E.	

9. Radioactive material shall only be stored and used at:

<u>Site Number</u>	<u>Location</u>
001	Richardson – 710 Presidential Drive
002	Richardson – 525 International Parkway

10. Each site shall maintain documents and records pertinent to the operations at that site. Copies of all documents and records required by this license shall be maintained for Agency review at **Site 002**.
11. The licensee shall comply with the provisions (as amended) of Title 25 Texas Administrative Code (TAC) §289.201, §289.202, §289.203, §289.204, §289.205, §289.251 and §289.252.
12. The individual designated to perform the functions of Radiation Safety Officer (RSO) for activities covered by this license is **Lorraine L. Hellinghausen**.
13. Radioactive material shall be used by, or under the direct supervision of, individuals designated by the RSO only after each worker has successfully completed a training course determined by the Agency as appropriate. Documentation verifying the successful completion of the training for each worker shall be maintained by the licensee for inspection by the Agency.
14. Sealed sources containing radioactive material shall not be opened.
15. The licensee shall conduct a physical inventory, at least every six months, to account for all sealed sources received and possessed under the license. The records of the inventories shall be maintained for inspection by the Agency for three years from the date of the inventory and shall include the radionuclide, number of curies, location of each source of radiation, the name of the individual making the inventory, and the date of the inventory.



Department of State Health Services

RADIOACTIVE MATERIAL LICENSE

LICENSE NUMBER	AMENDMENT NUMBER
L05519	09

16. The licensee shall conduct radiation contamination surveys of all radioactive material use and storage areas at intervals not to exceed 30 days or when it is suspected that a tritium light source has become compromised. Surveys shall be taken in the form of surface wipes using appropriate media for subsequent gas proportional counting or liquid scintillation counting. If analysis reveals the presence of radioactive contamination in excess of the limits contained in 25 TAC §289.202(ggg)(6), the licensee shall immediately enable contamination control actions and notify the agency of the analysis findings.
17. Except as specifically provided otherwise by this license, the licensee shall possess and use the radioactive material authorized by this license in accordance with statements, representations, and procedures contained in the following:

application dated March 1, 2012, and
letters dated **January 29, 2013** and **February 21, 2013**.


Title 25 TAC §289 (as amended) shall prevail over statements contained in the above documents unless such statements are more restrictive than the regulations.

JMK

FOR THE DEPARTMENT OF STATE HEALTH SERVICES

Date

February 22, 2013


Jason M Kelly, MPH, CPH, Program Coordinator
Advanced Technology Licensing Program



TEXAS DEPARTMENT OF STATE HEALTH SERVICES

DAVID L. LAKEY, M.D.
COMMISSIONER

P.O. Box 149347
Austin, Texas 78714-9347
1-888-963-7111
TTY: 1-800-735-2989
www.dshs.state.tx.us

February 22, 2013

TRUGLO INC
ATTN LORRAINE L HELLINGHAUSEN
525 INTERNATIONAL PKWY
RICHARDSON TX 75081

RE: License No. L05519
Application No. 10070


Dear Mr. Hellinghausen:

Enclosed is amendment number 09 to the above referenced license. This amendment closes Application No. 10070.

After you have completed the move, you will need to request termination of the former location. Prior to releasing the former location to unrestricted use, you must submit survey and sample results verifying compliance with the 25 Texas Administrative Code (TAC) §289.202(ccc)(2). Therefore, please conduct the required close-out survey and sample analysis and submit your results to the department, having the results keyed to a diagram illustrating the locations where the readings/wipes were obtained. In addition, please identify the make, model and last date of calibration of the radiation detection/analysis equipment used to take the assessment.

If I may be of assistance with your radiation safety program, please contact me at (512) 834-6688 extension 2205; or by electronic mail at jason.kelly@dshs.state.tx.us.

Sincerely,



Jason M Kelly, MPH, CPH, Program Coordinator
Advanced Technology Licensing Program
Radioactive Material Licensing Group

Enclosure(s)

bcc:file