

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Salahi Heart Clinic
1135 West University
Suite 355
Rochester Hills, Michigan 48095

REPORT NUMBER(S) 2013-01

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-37602

4. LICENSE NUMBER(S)

21-32672-01

5. DATE(S) OF INSPECTION

02/28/2013

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☒ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Bill C. Lin		3/5/13
BRANCH CHIEF	T. Bloomer		3/8/13

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Salahi Heart Clinic
1135 West University
Suite 355
Rochester Hills, Michigan 48095

REPORT NUMBER(S) 2013-01

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-37602

4. LICENSE NUMBER(S)

21-32672-01

5. DATE(S) OF INSPECTION

02/28/2013

6. INSPECTION PROCEDURES USED

87130

7. INSPECTION FOCUS AREAS

3.01-3.07

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)

02201

2. PRIORITY

5

3. LICENSEE CONTACT

Vince McCormick, RSO

4. TELEPHONE NUMBER

(734) 395-9323

☒ Main Office Inspection

Next Inspection Date: February, 2018

☐ Field Office Inspection

☐ Temporary Job Site Inspection

PROGRAM SCOPE

The licensee was a small cardiac diagnostic office authorized to use licensed material permitted by Title 10 Code of Federal Regulations 35.200. The nuclear medicine department was staffed with one technologist who performed approximately seven diagnostic cardiac procedures per day. The licensee is currently authorized by its NRC Materials License to perform diagnostic cardiac procedures at two locations. At the time of the inspection, the clinic only performs cardiac diagnostic procedures every other Thursday at the Rochester Hills location, and every other Monday at its Washington, Michigan location. The licensee received unit doses from the local nuclear pharmacy. The licensee's RSO/consultant audited the radiation program on a quarterly basis (last 12/20/2012).

Performance Observations

The inspection consisted of interviews with licensee personnel and a review of select records such as dosimetry, annual audits, survey instrument calibration, weekly wipes, and radiation safety training. The inspector also toured the facility and performed independent radiation surveys. Interviews conducted with the NMT revealed an adequate level of understanding of emergency and material handling procedures and techniques. Dose calibrator constancy checks, daily surveys, and package receipt were successfully demonstrated and observed by the inspector. The inspector observed two administrations of licensed material. The inspector closed out the previous violation of failure to perform weekly wipe surveys.

No violations of NRC requirements were identified.