

U.S. NUCLEAR REGULATORY COMMISSION**Date:** 3/20/2012 10:00AM**TELEPHONE CONVERSATION RECORD****Mail Control
or Report No(s).****License No(s).** 47-09772-02**Docket No(s).** 03003390**Name of Licensee:** Camden Clark Memorial Hospital Corporation**Name of Participant(s):** Daniel Berkley – RSO**Telephone No.** (304) 424-2744**Subject:** RAI concerning inspection findings
(NOTE: This will be used as the Documents Title in ADAMS)

In a telephone call on 3/20/12, I requested that the licensee look at the small volume of maximally exposed tissue outside the target in the post-plan isodose distribution, then compare that dose to the same volume of tissue in the pre-plan.

Action Required: None**Document Availability:**☒**Publicly Available**☐**Non-Publicly Available**☒**Non-Sensitive**☐**Non-Sensitive Copyright**☐**Sensitive- Proprietary**☐**Sensitive – Privacy Act/ PII**☐**Sensitive – Internal**☐**Sensitive – Security-Related****Immediate Release****Normal Release Date:** 04/15/2013**Delayed Release Date:****SUNSI Review Completed By:** TLWeidner / RA /