



TEST-CON INCORPORATED

16 East Franklin Street

P. O. Box 3116

Danbury, CT 06813-3116

Phone: (203) 748-3012 Fax: (203) 778-0633

E-Mail: Service@test-con.com Web-Site: www.test-con.com

Construction Material Testing & Inspection - Geotechnical Services

April 9, 2013

Sattar Lodhi
2100 Renaissance Blvd
Suite 100
Att. DNMS
King of Prussia, PA 19406
Fax: 610-337-5269

Q2

03038501

06-31458-01

Re: Nuclear Gauge Transfer Seaman C-300 Serial # 21182

Mr: Lodhi

We have received the gauge listed in the reference line from DeCarlo & Doll Inc. I am attaching a copy of the current Leak Test Certificate for your records.

Sincerely

A handwritten signature in black ink, appearing to be 'Chin Okwuka'.

Chin Okwuka
RSO

Attachments

Sent by Fax.

REC'D IN LAT 04-09-13

580396
NMSS/RGN1 MATERIALS-002

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA
info@seamannuclear.com, Tel 414-762-5100 Fax 414-762-5106

Leak Test Certificate

November 27, 2012

A leak test has been performed on meter, serial number **21182**, a model C-300, containing the radionuclide Ra 226.

Owned By: DeCarlo & Doll Inc.
89 Colony St
Meriden, CT 06451

Date Sample Collected: 11/6/12
Collected By: K. Rekrut
Date Sample Analyzed: 11/27/12
Analyzed By: T. Lindemann

Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

LEAK TEST DUE: 5/6/13



Scott C. Seaman
President & Radiation Safety Officer, Seaman Nuclear Corporation