

From: [Thompson, Thomas](#)
To: analyze@baronconsulting.com
Subject: Renewal of your NRC license 06-17669-01
Date: Wednesday, April 24, 2013 5:44:00 PM

I reviewed your April 9, 2013 letter in which you included leak test information for the GC's , however, I do not see leak tests for the two specifically licensed GC's serial number 5182 & F2981. Am I missing something?