

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED Oliver W. Caminos, M.D. d/b/a Pittsburgh Cardiovascular Institute and Medical Health Services Management, LP 200 James Place Monroeville PA 15146 REPORT NUMBER 2013-001		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 2100 Renaissance Blvd, Suite 100 King of Prussia, Pennsylvania 19406-2713	
3. DOCKET NUMBER(S) 150-00037	4. LICENSE NUMBER(S) PA-0449	5. DATE(S) OF INSPECTION February 27, 2013	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records. Interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☐ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s) and corrective action(s):




- ☒ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)

10 CFR 20.1802 requires, in part, that the licensee shall control and maintain constant surveillance of licensed material that is in a controlled area and that is not in storage. Contrary to the above, on February 27, 2013, the licensee did not control and maintain constant surveillance of licensed material that was in a controlled area and that was not in storage. Specifically, the technician left 175 mCi of Tc-99m in the camera room unlocked and unsupervised. This is a Severity Level IV violation (Section 6.7)

The licensee committed to securing the licensed material when not attended.

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Alfred A. Caminos		4/15/2013
NRC INSPECTOR	Robin Elliott		4/15/2013
BRANCH CHIEF	James P. Dwyer		4/24/13

SUNSI Review Completed By: / RA / RElliott



Public



Non-Sensitive