



UNITED STATES
NUCLEAR REGULATORY COMMISSION
WASHINGTON, DC 20555 - 0001

May 15, 2013

Jerry Szpak
Director of Technology
Met One Instruments, Inc.
1600 Washington Blvd.
Grants Pass, Oregon 97526

Mail Control No. 579474

SUBJECT: REQUEST FOR ADDITIONAL INFORMATION - AMENDMENT TO
EXEMPT DISTRIBUTION MATERIALS LICENSE

Dear Mr. Szpak,

This letter is in response to your amendment request dated November 20, 2012, to include the BAM 1022 Beta Attenuation Monitor to License No. 36-23875-01E. In reviewing your request, we find that additional information is required to complete our review. In the enclosure of this letter, we have summarized the issues not addressed in your letter.

Please submit the requested information within 30 days of the date of this letter. If we have not received complete information within 30 days of the date of this letter, we will consider your application as having been abandoned by you. This is without prejudice to the submission of a complete application.

In accordance with 10 CFR 2.390 of NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in NRC's Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). ADAMS is accessible from the NRC web site at <http://www.nrc.gov/NRC/ADAMS/index.html> (the Public Electronic Reading Room).

Any correspondence regarding your amendment application should reference the control number specified above.

J. Szpak

– 2 –

If you have questions regarding the amendment to your exempt-distribution materials license please contact me at (301) 415-6202, or by email at Katie.Wagner@nrc.gov.

Sincerely,

/RA/

Katie Wagner

Licensing Branch

Division of Materials Safety and

State Agreements

Office of Federal and State Materials and

Environmental Management Programs

Enclosure: As stated

Docket No. 030-35578

License No. 36-23875-01E

J. Szpak

– 2 –

If you have questions regarding the amendment to your exempt-distribution materials license please contact me at (301) 415-6202, or by email at Katie.Wagner@nrc.gov.

Sincerely,

/RA/

Katie Wagner
Licensing Branch
Division of Materials Safety and
State Agreements
Office of Federal and State Materials and
Environmental Management Programs

Enclosure: As stated

Docket No. 030-35578
License No. 36-23875-01E

DISTRIBUTION: MSSA TA RJones,LFARB RIV/DNMS MSSA r/f

ML12334A195 (Pkg.)
ML13135A201

OFFICE	FSME/MSSA/LB	C	FSME/MSSA/LB	C	FSME/MSSA/LB	C		
NAME	KWagner		SXu		MKotzalas/THerrera for:			
DATE	05/15/ 2013		05/15/ 2013		05/15/ 2013			

OFFICIAL RECORD COPY

ENCLOSURE

Questions pertaining to your Exempt-Distribution Materials License:

1. Section 32.26(b)(8), 10 CFR 32, requires the applicant to submit information that should include total quantity of byproduct material expected to be distributed in the product annually. Please submit information on the quantity of byproduct material expected to be distributed annually as part of BAM 1022 Beta Attenuation Monitors.
2. Section 32.29(b)(3), 10 CFR 32, states that each licensee under 32.26 shall label or mark its point-of-sale package so that the external surface of the point-of-sale package has a legible, readily visible label or marking containing information described in 32.29(b)(3)(i-iii), 10 CFR 32. Please identify any differences in how the BAM 1022 point-of-sale packaging will be labeled from the models currently listed on your exempt-distribution materials license.
3. In your response dated 13 March 2013, you stated that "The BAM 1022 meets the dose limits of 10 CFR 32.27." Please provide a rationale for how this conclusion was reached (e.g. provide a calculation or other information).