

From: [Bensen, Debbie](#)
To: [Jankovich, John](#)
Subject: FW: Changes to the Flexitron unit
Date: Wednesday, October 30, 2013 10:32:46 AM
Attachments: [image36bf8c.PNG](#)
[GA-0497-D-115-S.pdf](#)
[Flexitron system label.pdf](#)
[label TDU flexitron 110V.PDF](#)
[Flexitron configuration label.pdf](#)
Importance: High

Dear John,

Some time ago I sent an email to you requesting guidance. Nucletron B.V. in Veenendaal has changed the labels for the Flexitron, see attached. They currently have devices ready to be shipped but I have not heard back from you whether they can be shipped while we are waiting for the SS&D. Can you please look at the attached and get back to me? I am under a tremendous amount of pressure from Company management.

Thank you,
Debbie

Debra Bensen | Radiation Safety Officer
Elekta
400 Perimeter Center Terrace, Suite 50
Atlanta, GA 30346, United States
Office: +1-770-670-2518 | Mobile: +1-410-913-3162 | Fax: +1-443-769-1577
debbie.bensen@elekta.com | www.elekta.com



From: Bensen, Debbie
Sent: Thursday, October 03, 2013 3:22 PM
To: 'john.jonkovich@nrc.gov'
Subject: Changes to the Flexitron unit

Dear John,

I hate to be a pain, but working with my mother company in the Netherlands is sometimes a little trying. It seems they have changed the configuration of the Flexitron in that it can now be purchased with a 10 or 20 channel configuration in addition to the original 40 channel configuration that is registered. They have also updated the software and user's manual as well as changed the labels. (New labels attached) I do need to submit an amendment for these changes (correct me if I am wrong) but are these changes enough to stop delivery and installation while the SS&D is being amended? Please advice.

Thank you,
Debbie

Please consider the environment before printing this e-mail.

The contents of this e-mail message (including any attachments) are confidential to and are intended to be conveyed for the use of the recipient to whom it is addressed only. If you receive this transmission in error, please notify the sender of this immediately and delete the message from your system. Any distribution, reproduction or use of this message by someone other than recipient is not authorized and may be unlawful.