

Br. 1

**PriMed**  
**PHYSICIANS**

United For Patient Care  
112 Quarry Road, Trumbull, CT 06611  
tel. 230-333-8800 fax 203-335-8275

To: USNRC Region I Office  
Licensing Assistance Team  
2100 Renaissance Blvd, Suite 100  
King of Prussia, PA 19406-2713

November 25, 2013

RE: Amendment request- removal of the Bridgeport office site to NRC Materials License  
06-30537-01. Docket # 030-35233

REC'D 12/03/13 AM 07:34

To Whom It May Concern:

This letter serves to request a license amendment for PriMed, LLC, d/b/a Connecticut Heart & Vascular Center, PC. We have stopped the radioactive material delivery and patient imaging at the 2979 Main Street, Bridgeport office. The nuclear cardiac imaging equipment has now been relocated to the new office in Trumbull, CT.

November 20, 2013 was the last day of radioactive material delivery for patient imaging at the Bridgeport office. Our "end of day" radiation level surveys and wipe tests for radioactive contamination (refer to following page) were conducted and found satisfactory. The reference sources (for dose calibrator QA and camera flood source QA) were brought to the new office and securely stored in the Hot Lab.

Please delete the previous work address of Bridgeport, CT from our license, so we may relinquish use of this space to new tenants. If you should have any questions or require additional information regarding our nuclear cardiology imaging office please, contact me at 203-333-8800.

We appreciate your efforts to process this amendment and we look forward to a continued safe and effective materials program. Thank you.

Respectfully submitted,

  
Robert Sackstein, MD, RSO

The following data summarizes the final survey for ambient radiation (mR/hr) levels and removable contamination (dpm) from selected areas of the closed office. There were no areas demonstrating contamination and the final survey results were:

<u>Location / Area</u>	<u>mR/hr</u>	<u>dpm</u>
Imaging Room		
patient table	≤ 0.02	15.5
acquisition station	≤ 0.02	12.3
processing station	≤ 0.02	14.9
Hot Lab		
Prep area	≤ 0.03	30.4
delivery area	≤ 0.02	20.2
injection area	≤ 0.03	18.0
Prep Room 1		
counter top	≤ 0.04	30.6
patient table	≤ 0.02	28.9
Stress Room 1		
counter top	≤ 0.02	15.4
treadmill	≤ 0.02	12.8
Stress Room 2		
counter top	≤ 0.02	34.2
treadmill	≤ 0.02	41.2
Echo Room 1	≤ 0.02	23.7
Echo Room 2	≤ 0.02	16.4
Bathroom	≤ 0.02	19.8

The background radiation levels were: 0.01 - 0.03 mR/hr (ambient survey), and, the removable contamination is NET dpm (contamination wipe). The survey instruments used were: Ludlum Model 14C GM Survey Meter, and Capintec Caprac shielded NaI crystal with a well.



This is to acknowledge the receipt of your letter/application dated

11/25/13, and to inform you that the initial processing which includes an administrative review has been performed.

☒ Amendment (06-30537-01)  
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 582639.  
When calling to inquire about this action, please refer to this control number.  
You may call us on (610) 337-5398, or 337-5260.