




UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
1600 EAST LAMAR BLVD
ARLINGTON, TEXAS 76011-4511

EMAIL



Name: Alan M. Vetter License: 50-23204-01
Organization: Shannon and Wilson, Inc. Docket: 030-20360
Phone: 907-479-0600 Control: 581290
E-mail Address: amv@shanwil.com
From: Jacqueline D. Cook
Date: December 21, 2013
Subject: Application dated June 5, 2013 and Response to Request for Additional Information dated November 26, 2013 for License Renewal
Pages: 2

Mr. Vetter:

Per your application dated June 5, 2013, and response to request for additional information dated November 26, 2013, the items on the next page are deficiencies which require your response. **Please respond to this e-mail by Monday, January 6, 2014.** Our fax number is (817) 200-1263. You may respond by e-mail in pdf format if you'd like. My email address is Jackie.Cook@nrc.gov. When responding to this e-mail, please include the license, docket and control numbers located at the top of this page.

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

/RA/
Jacqueline D. Cook
Senior Health Physicist

PUBLIC

- ☐ Immediate Release
☒ Normal Release

NON-PUBLIC

- ☐ A-1 Sensitive-Security Related
☐ A-7 Sensitive Internal
☐ Other: _____

Reviewer: JDC Date: 12/21/13

1. Please expand your response to Note #3 to include "security."
2. Although you addressed non-routine maintenance procedures for Troxler Model 3400 series, please address non-routine maintenance procedures for Troxler Model 3241-C. Please clarify if these procedures will be performed in-house or will the gauge be sent to the manufacturer or other person authorized by NRC or an Agreement State to perform non-routine maintenance or repair operations that require the removal of the source or source rod from the gauge. If you request to perform non-routine maintenance on Troxler Model 3241-C in-house, please submit the information outlined in Appendix G of NUREG-1556, Vol. 1. (<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v1/r1/>).

PUBLIC

- ☒ Immediate Release
☒ Normal Release

NON-PUBLIC

- ☐ A.3 Sensitive-Security Related
☐ A.4 Sensitive Internal
☐ Other

Reviewed: Date: 12/20/13