

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Genesys Hurley Cancer Institute  
302 Kensington Avenue  
Flint, Michigan 48503

REPORT NUMBER(S) 2014-01

2. NRC/REGIONAL OFFICE

Region III  
U. S. Nuclear Regulatory Commission  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-36106

4. LICENSE NUMBER(S)

21-32322-01

5. DATE(S) OF INSPECTION

1/30/14

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Bill C. Lin		1/30/14
BRANCH CHIEF	Daniel T. McCremas		2/6/14

## Docket File Information

## SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:  Genesys Hurley Cancer Institute 302 Kensington Avenue Flint, Michigan 48503  REPORT NUMBER(S) 2014-01		2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352	
3. DOCKET NUMBER(S)  030-36106	4. LICENSE NUMBER(S)  21-32322-01	5. DATE(S) OF INSPECTION  01/30/2014	
6. INSPECTION PROCEDURES USED  87131, 87132	7. INSPECTION FOCUS AREAS  ALL		

## SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)  02230	2. PRIORITY  2	3. LICENSEE CONTACT  Ahmed Akl, M.D., RSO	4. TELEPHONE NUMBER  (810) 762-8490
<input checked="" type="checkbox"/> Main Office Inspection <input type="checkbox"/> Field Office Inspection <input type="checkbox"/> Temporary Job Site Inspection			
Next Inspection Date: January, 2016			

## PROGRAM SCOPE

This was a Routine Inspection of a free-standing cancer treatment center with authorization to use materials in Section 35.300 and 35.600. The radiation oncology department was staffed with three authorized medical physicist (AMP), and three authorized physician users. The licensee administered thyroid carcinoma treatments, Zevalin treatments, Radium 223 prostate treatments, and HDR treatments. All HDR patient treatments were administered by the attending radiation oncologist and the AMP. Service, maintenance, and source exchanges were performed by the HDR device manufacturer. The licensee hired an outside consultant to audit the licensee's Section 35.300 program.

## Performance Observation

This inspection consisted of interviews with select licensee personnel; a review of select records; a tour of the radiation oncology department; and independent measurements. The inspector observed the licensee staff performed the quality assurance check on the HDR, reviewed patient's written directives and the treatment plan, and interviewed the attending physician and AMP.

No violations of NRC regulatory requirements was identified.