

Junod, Rebecca

From: Seeley, Shawn
Sent: Tuesday, February 04, 2014 7:59 AM
To: kwisell@echn.org
Cc: Weidner, Tara
Subject: RE: Manchester Memorial Renewal MC582636

Mr. Wisell,

After sending this to you I noticed an error with the request for item #2. The reference to the HDR room should have been a request for a diagram of the facility/room utilized for the I-125 implants. I apologize for the confusion and any inconvenience this may have caused.

Shawn

Shawn W. Seeley, Health Physicist
USNRC-RI, Division of Nuclear Materials Safety
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713

Office: 610-337-5102 Fax: 610-337-5269

From: Seeley, Shawn
Sent: Friday, January 31, 2014 7:44 AM
To: 'kwisell@echn.org'
Cc: Weidner, Tara
Subject: FW: Manchester Memorial Renewal MC582636

Mr. Wisell,

This email is in regard to your renewal application dated November 27, 2013 for license number 06-03413-01. Please refer to Mail Control (MC) 582636 when responding. In order to complete our review of your application, we need the following additional information.

- 1: Your application contains two different addresses for the place of use, i.e 70 Haynes Street and 71 Haynes Street. Please confirm which is the address you wish to use.
- 2: Your application did not contain a facility diagram for the nuclear medicine or HDR areas. Please submit a facility diagram of these two areas and include a description of the areas surrounding the areas as well as the areas above and/or below. Please refer to section 8.16 of Nureg-1556, Vol. 9, Rev. 2 for guidance.
- 3: Your application makes reference to a "general storage room" within the hot lab area. Please provide additional details on the security and accessibility of this room.
- 4: We would like to bring to your attention two incorrect references in your procedures:
 - A: In Item 9.5, Other Equipment and Facilities, there is a reference to 20.105(b) of 10 CFR 20. Please note that this citation is incorrect and should be 20.1301.
 - B: Item 10.5, Spill Procedures, contains a reference that is not necessary. Please remove the reference to 10 CFR 20.1301 as it not necessary.

Please submit the information requested in items 1 through 3 above via facsimile or hard copy to my attention. Please include the reference to MC582636. Also include a cover letter signed by a member of management. If you have any questions, please let me know.

Thank you.

Shawn

Shawn W. Seeley, Health Physicist
USNRC-RI, Division of Nuclear Materials Safety
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713

Office: 610-337-5102 Fax: 610-337-5269