

Honeywell International

June 18, 2007

William Wright
Arizona Radiation Regulatory Agency
4814 South 40th Street
Phoenix, Arizona 85040

Subject: Amendments to Sealed Source and Device Registration Certificates.

Dear Mr. Wright:

This letter is a request to amend the Sealed Source and Device Registry certificates issued to Honeywell International. We are requesting that AZ0501D105B, AZ0501D106B, AZ0501D107B, and AZ0501D108B, be amended to add two new manufacturing sites and an additional location for distribution.

Currently the SS& D certificates list the Distributor/Manufacturer as follows:

<u>DISTRIBUTOR/MANUFACTURER:</u>	Honeywell-International Radiological Operations (G-20) 2500 West Union Hills Drive Phoenix, Arizona 85027-5139 (602) 313-3330
----------------------------------	---

Amend to read as follows:

MANUFACTURER:

Honeywell International Radiological Operations (G-20) 2500 West Union Hills Drive Phoenix, Arizona 85027-5139 (602) 313-3330	Honeywell Measurex, Ltd Waterford Industrial Estate Waterford Ireland +(353) 51 372 151	Honeywell ASCa, Inc 500 Brooksbank Avenue North Vancouver, BC V7J3S4 Canada (604) 982-3607
---	---	--

<u>DISTRIBUTOR:</u>	Honeywell-International Radiological Operations (G-20) 2500 West Union Hills Drive Phoenix, Arizona 85027-5139 (602) 313-3330	Honeywell International 3079 Premiere Parkway Suite 100 Duluth, Georgia 30097 (770) 689-0197
---------------------	---	--

Further we commit that all Stainless Steel parts of our gauges are constructed using Series 300 stainless steel and that all Tungsten parts are constructed using tungsten that contains at least ninety percent tungsten with the other ten percent being a combination of iron, copper and nickel.

3079 Premiere Parkway
Suite 100
Duluth, Georgia 30097

Honeywell International
June 18, 2007
Page 2 of 2

This change will ensure that all the devices are consistent in their construction and that any references to specific manufacturers and brand names of stainless steel and tungsten (i.e. Mallory) are superseded. The change will not change any of the radiation protection characteristics of the devices.

Honeywell will ensure that each of the gauges will continue to comply with our existing commitments (maximum dose rates, labeling, ANSI classification, etc.) as stated in the applicable registry certificates.

Thank you for your assistance in this matter. If you require additional information to complete your review please don't hesitate to contact us at (770) 689-0197 or by email at rodriquez.harrell@honeywell.com.

Sincerely,



Gary L. Caines
Radiological Operations Program Manager

Enclosures: AZ0501D105
AZ0501D106
AZ0501D107
AZ0501D108