

## **Nimmo, Elsa (CA35-x2)**

**From:** Nimmo, Elsa (CA35-x2)  
**Sent:** Thursday, June 11, 1998 4:49 PM  
**To:** 'Wesley, David (RHB)'  
**Cc:** French, Wes (CA35-x2); Axelrod, Steve (CA35-x2); Frohardt, Allen (CA35-x2); Wiley, Scott (CA35-x2); Seright, Bill (CA35-x2); Lisa Burns; Nell Lyon  
**Subject:** RE: Modification of drawings

David,

Thanks, that is good news!

The piece part drawing in question does change from its current Rev. B to Rev. C with the implementation of this change. There will be no change in the assembly dwg rev level.

Thanks very much for this quick response!

Elsa

-----  
**From:** Wesley, David (RHB)[SMTP:dwesley@rhb.dhs.cahwnet.gov]  
**Sent:** Thursday, June 11, 1998 4:00 PM  
**To:** 'Nimmo, Elsa (CA35)'  
**Subject:** Modification of drawings

Hi Elsa,

Your voice message got cut off, but I got the bulk of the message and Tom Schell showed me the fax.

After a quick review, it is apparent that the change will not affect the shielding or shutter operation (other than extending the lifetime of the beam-off detecting switch), and therefore the Department of Health Services approves the revision.

I don't believe we need to make any changes to the certificate, unless it changes the assembly drawing revision letter (let me know if it does). Otherwise, we will shred the last page of the fax (engineering drawing of the Source holder cam wheel which is marked "Confidential") and put the rest of the fax in our file along with a copy of this e-mail.

Let me know if there is anything else we can do.

~ David Wesley  
(916) 445-1884