

Junod, Rebecca

From: Seeley, Shawn
Sent: Tuesday, March 04, 2014 6:51 AM
To: nmip.bruce@gmail.com
Subject: Request For Additional Information Lic#47-30849-01, MC582723

Mr. Morton,

This email is in response to the renewal application for Cardiovascular Consultants, PLLC, located in Charleston, West Virginia. License number 47-30849-01, docket number 030-36436, and mail control number 582723. Please refer to this mail control number when replying. In order to continue our review, we need the following additional or clarifying information submitted:

- 1: The delegation of authority appears to have been signed by the RSO and not the management representative. Please submit a delegation and acceptance of authority signed by the management representative.
- 2: The facility diagrams that were submitted were not legible. Please provide updated diagrams and include a description of the areas adjacent and immediately above or below the restricted areas.
- 3: Please confirm that "you reserve the right to upgrade your survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used."
- 4: Please note that the hazardous material training program is missing the "security awareness" topic. Please confirm that this topic will be added to the training program.

The above information can be faxed to me at 610-337-5269. Thank you for your cooperation.

Shawn

Shawn W. Seeley, Health Physicist
USNRC-RI, Division of Nuclear Materials Safety
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713

Office: 610-337-5102 Fax: 610-337-5269