

From: [Gaskins, Farrah](mailto:Gaskins.Farrah)
To: JerryN@JJAconstruction.com
Subject: Request for Additional Information
Date: Wednesday, March 19, 2014 2:47:00 PM

Dear Mr. Naimoli,

This is in reference to your application dated January 27, 2013 requesting to renew Nuclear Regulatory Commission License No. 37-28163-01. In order to continue our review, we will need the following additional information:

1. In item 3 of your application, you listed an address in Pennsylvania as the address where licensed materials will be used or stored. Since Pennsylvania is an Agreement State, the NRC will only authorize use of licensed material at temporary job sites anywhere NRC has Jurisdiction. Please specify that you are requesting use of material at temporary jobsites.
2. Please confirm that "Before using licensed materials, authorized users will have successfully completed one of the training courses described in Criteria in the section entitled "Training for Individuals Working In or Frequenting restricted Areas" in NUREG-1556, Volume 1, Revision 1, dated November 2001.
3. Please confirm that you will either possess and use, or have access to and use, a radiation survey meter that meets the Criteria in the section entitled "radiations Safety Program – Instruments" in NUREG-1556, Volume 1, Revision 1, dated November 2001.

The NRC's Safety Culture Policy Statement became effective in June 2011. While a policy statement and not a regulation, it sets forth the agency's *expectations* for individuals and organizations to establish and maintain a positive safety culture. You can access the policy statement and supporting material that may benefit your organization on NRC's safety culture Web site at <http://www.nrc.gov/about-nrc/regulatory/enforcement/safety-culture.html>. We

strongly encourage you to review this material and adapt it to your particular needs in order to develop and maintain a positive safety culture as you engage in NRC-regulated activities.

In order to continue prompt review of your application, we request that you submit your response to this email within 30 calendar days from the date of this email. Your response must be signed by a member of your management and either faxed (number provided below) or sent electronically as a pdf file. Please reply to my attention at the Region I Office and refer to Mail Control No. 583169. If you have any technical questions regarding this deficiency email, please call me at (610) 337-5143.

Farrah C. Gaskins

Health Physicist
U.S. NRC, Region 1
2100 Renaissance Blvd, Suite 100

King of Prussia, PA 19406
(610) 337-5143 office
(610) 337-5269 fax