



RECEIVED

JUN 27 1995

RADIOACTIVE MATERIALS PROGRAM

June 26, 1995

Cynthia Sanders
Radiological Health Specialist, Senior
Radioactive Materials Program
Georgia Department of Natural Resources
4244 International Parkway
Suite 114
Atlanta, GA 30354

Dear Ms. Sanders:

In answer to your letter of June 21, 1995, we have the following comments.

- A. Electrical Testing Laboratories (ETL) does not approve the design. Their function is to insure that approved electrical components are used. Since we have incorporated an additional motor and wiring as part of the filter modification, we were obliged to have the components submitted for approval to insure continued ETL approval of the complete unit.
- B. All units which are installed after July 1, 1995 will have the modification made to them prior to customer turnover. Our intention is to complete the modification to the 20 operational sites by December 31, 1995.
- C. As per your letter, a request for amendment of the device registry will be made under a separate letter.

Regards,


Martin Knotts
Project Manager