

Georgia Department of Natural Resources

4244 International Parkway, Suite 14, Atlanta, Georgia 30354

Lonice C. Barrett, Commissioner

Environmental Protection Division

Harold F. Reheis, Director

(404) 362-2675

June 21, 1995

Martin Knotts
Director of Technical Services
Elekta Radiosurgery, Inc.
8 Executive Park West, Suite 809
Atlanta, Georgia 30329

Dear Mr. Knotts:

Thank you for your letter dated May 10, 1995 to Mr. Thomas Hill regarding the gamma knife incident recommendations. We have reviewed your corrective actions and do require the following additional information.

- A. Please provide us with the results from Electric Testing Lab (ETL) on the new filtering system. Did they approve the new design?
- B. Submit a proposed schedule for installation of the new filtering system at all of your operational sites.
- C. Since modifications have taken place on the filtering system to the bed and check valves will be installed, you are now required to amend the device registry to incorporate this change.

If you have any questions, please feel free to contact this office at (404) 362-2675.

Sincerely,



Cynthia Sanders
Radiological Health Specialist, Senior
Radioactive Materials Program