

HANSELL & POST

A PARTNERSHIP INCLUDING PROFESSIONAL CORPORATIONS

56 PERIMETER CENTER EAST, N.E.
FIFTH FLOOR
ATLANTA, GEORGIA 30346-2283
TELEPHONE 404 399-1600
TELECOPIER 404 394-4637
TELEX 80-4455

THIRTY-THIRD FLOOR FIRST ATLANTA TOWER
ATLANTA, GEORGIA 30383-3101

1915 "I" STREET, N.W.
FIFTH FLOOR
WASHINGTON, D.C. 20006

November 8, 1985

VIA HAND DELIVERY

Mr. Tom Hill
Georgia Radiological Health Section
878 Peachtree Street, Room 600
Atlanta, Georgia 30309

Re: Application for Nuclear Device
Registration---Leksell Gamma System

RECEIVED

NOV 8 1985

RADIOLOGICAL
HEALTH UNIT

Dear Mr. Hill:

Pursuant to our telephone conversation of October 30, 1985, this is to confirm the following information concerning the Leksell Gamma System Model No. 23016:

1. Consistent with Section 4.2.2(c) of NCRP Report No. 33 and as indicated on page 2 of the pending Application for Nuclear Device Registration, the manufacturer recommends that the user wipe or smear test the outer (convex) surface of the collimator helmets every six months. Part G-1 of the User's Manual contains the same recommendation. The first sentence of Part G.3 of the User's Manual will be amended to read:

"Helmets are to be cleaned of dust and dirt inside and outside by dry wiping once every six months, immediately following the leak test (see G.1)."

2. The "cover at the back of the helmet" referred to under the heading "COVER OPEN" on page 14 of our September 17, 1985 responses to your questions and in Part D.3.9 of the User's Manual, is fully described in Drawing No. 13030 found in Part J.8 of the User's Manual and is identified as Item No. 24 in Drawing No. 13029 found in the same part of the User's Manual. Reproductions of Drawings 13029 and 13030 are enclosed for convenient reference.

Mr. Tom Hill
Page Two
November 8, 1985

If you have any further questions, please do not hesitate to call.

Sincerely,

A handwritten signature in black ink, appearing to read 'J. Maiers', written over a horizontal line.

John D. Maiers

JDM:wjh

Enclosures as stated