



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

TELEFAX TRANSMITTAL

DATE March 12, 2014

NUMBER OF PAGES 2

SEND TO Brian Buter, CNMT, Lead Nuclear Medicine Technologist, NRC
License 21-16737-01, Memorial Medical Center of West Michigan

LOCATION Ludington, Michigan

FAX NUMBER (231) 845-3611

☒ **VERIFY BY CALLING**
SARA CONFIRMED FAX RECEIVED.

FROM: Bill Reichhold
(Sender)

TELEPHONE NUMBER (630) 829-9839

FAX NUMBER (630) 515-1078

If you do not receive the complete fax transmittal, please contact the sender as soon as possible at the telephone number provided above.

MESSAGE See accompanying documents.

NOTICE

This message is intended only for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, or exempt from disclosure under applicable law. If the reader of this message is not the intended recipient or the employee responsible for delivering the message to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. If you received this communication in error, please notify the sender immediately by telephone and return the original to the above address, by U.S. Mail. Thank You.

The following additional information is needed to review your request to add a new imaging room and cardiac "stress" rooms.

1. Please put the address of the facility on the facility diagram.
2. Please specify the scale of the facility diagram, such as 1 inch = 1 foot. OR Specify the room dimensions (such as 14 feet by 12 feet) where radionuclides are used or stored.
3. Please specify the room numbers where radionuclides are used or stored. If there are no room numbers, please state so.
4. If you are using PET radionuclides, specify the location of the "quiet room". Also specify any additional shielding, remote handling devices, and confirm that you will perform surveys to ensure that you will not exceed the dose limits specified in 10 CFR 20.1301. If you are not using PET radionuclides, please state so.
5. Please specify what is above (such as roof) and below (such as basement) areas where radionuclides are used or stored.

Please send a facsimile (630- 515-1078) of your response to the above within 7 days and state, Response to Control 582978. Please include a cover letter on company letterhead, dated and signed (signed by an individual who is authorized to sign official documents on behalf of the licensee) with your response letter. Please call me at 630-829-9839 if you have any questions.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this facsimile and the attached documents will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

From the desk of:



Bill Reichhold