



Region I Office
Division of Nuclear Materials Safety
2100 Renaissance Boulevard
Suite 100
King of Prussia, PA 19406-2713
(610)337-5000

Telephone Conversation Record

Date: March 24, 2014

License No.: 41-25219-01

Docket No.(no hyphens): 03032925

Mail Control/Report No.: N/A

Licensee Name: Tennessee Valley Authority

Participant(s) Name/Title: Lee Miller/Project Manager

Work Telephone No.: 423-751-3197

Business Cellphone No.: None

NRC Representative Name/Title: Steven R. Courtemanche/Health Physicist

Subject: Additional Information for NMED 110349

(This is the title that will be used in ADAMS)

Discussion: INL requested information concerning whether the gauge had been repaired which was not covered in the licensee's LER (2011-014) dated August 12, 2011 (ML112360496). (3/21/2014, 09:00) I was informed by the receptionist that Joseph Shea, Vice President, Nuclear Licensing, (423-751-6887) was on the phone with another representative of the NRC and would return my call. (11:15) I was informed by the receptionist that Mr. Shea was still unavailable but she would get back in touch with me if he could not contact me today. I provided the receptionist with information on the purpose of the phone call so an alternate could be found if necessary. (14:00) Mr. Miller returned the phone call on behalf of Mr. Shea. I asked Mr. Miller to provide information about the repairs to the gauge with the stuck shutter described in the August 12, 2011, report. Mr. Miller stated that the area was posted with a "Danger" tag and the manual actuator for the shutter mechanism was locked. Mr. Miller further stated that the representative of the manufacturer had the gauge sent back to the manufacturer on August 5, 2011, for repairs. The repairs included a cleaning of the gauge, removal of the source holder, insertion of the old source in a new source holder, and placement of the new source holder in the gauge. The gauge was placed back in service on January 19, 2012.

Action Required: Document for next inspection. Close out NMED 110349.

SUNSI REVIEW

Document Availability: ☒ Public or ☐ Non-Public

Document Sensitivity:

(select "1" value to the right)

<input checked="" type="checkbox"/>	Non-Sensitive	<input type="checkbox"/>	MD 3.4 Non-Public B.1 (Non-Sensitive)
<input type="checkbox"/>	Non-Sensitive Copyright	<input type="checkbox"/>	MD 3.4 Non-Public A.3 (Sensitive Security Related
<input type="checkbox"/>	Sensitive – Proprietary	<input type="checkbox"/>	MD 3.4 Non-Public A.4 (Sensitive Proprietary
<input type="checkbox"/>	Sensitive – Privacy Act (includes Personally Identifiable Information (PII))	<input type="checkbox"/>	MD 3.4 Non-Public A.5 (Sensitive-Privacy Act)
<input type="checkbox"/>	Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.)	<input type="checkbox"/>	MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc)
<input type="checkbox"/>	Sensitive – Security-Related-Periodic Review Required	<input type="checkbox"/>	MD 3.4 Non-Public A.7 (Internal)

SUNSI Review Completed by: Steven R. Courtemanche