

MEMO

To: File
From: Eric Jameson *EJ*
Subject: SS&D Registry amendment, GA 269-D-102-S
Date: August 9, 1996

Summary of Telephone Conversation with Richard Grome on Friday, August 9, 10:30 am.

- Question regarding deleted item in Installation Test Protocol has been answered. Item to be performed was actually written twice in prior Functional Test.
- To improve clarity, the revision on page 8, paragraph 3 will be changed to read:

Important: A treatment stop cycle is initiated should a main power failure occur during a treatment. The batteries will continue to power the entire system including all function indicators (i.e., lamps). The operator should (continuation of initial submittal).

A copy of this change was faxed to Elekta later that morning.