



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION III
2443 WARRENVILLE RD. SUITE 210
LISLE, IL 60532-4352

MAR 19 2014

Andrea D. Browne, Ph.D.
Radiation Safety Officer
Community Health Network, Inc.
1500 N. Ritter Avenue
Indianapolis, IN 46219

Dear Dr. Browne:

We have reviewed your license renewal request dated December 23, 2013. As discussed with you by telephone on March 19, 2014, please provide the following.


- 1) With regard to the use of 10 CFR 35.400 material, for each radionuclide please provide the name, the sealed source models, and the possession limit.
- 2) With regard to the use of 10 CFR 35.500 material, for each radionuclide please provide the name, the sealed source models, and the possession limit.
- 3) The current license listed the possession limit for Iotrex permitted by 10 CFR 35.1000 as 8 Ci; however, your renewal application listed only 8 mCi. Please confirm the correct quantity.
- 4) You requested to name Ms. Ana Mihail as an Authorized Medical Physicist for a Strontium-90 ophthalmic source. Please provide training information to support your request (i.e., a completion of Form 313(AMP) with regard to the Strontium-90 ophthalmic training).
- 5) You requested the use HDRs permitted by 10 CFR 35.600. Based on NUREG-1556, Volume 9, Revision 2, please provide a statement that the licensee is providing the procedures required by 10 CFR 35.643 with regard to the HDR's calibration and use.
- 6) Except the new facility at the South campus that have not yet listed on your license, please provide a copy of information related to the shielding of rooms that use the HDRs and PETs and other information describes in Item 9, "Other Equipment and Facilities", of Appendix C to NUREG-1556, Volume 9, Revision 2. In addition, if the locations of use have not been changed since the last license amendment, please confirm. If there were changes, please describe.

NUREG-1556, Volume 9, Rev. 2 dated January 2008 provides the program specific guidance about medical use licenses. NUREG-1556, Volume 9 is available on the NRC website at <http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>.

To continue review of your request, we request that you submit your response to this letter by April 2, 2014. Your response must be **dated** and **signed** by authorized personnel and referred to Control Number 582780. If you have any questions or require clarification on any of the information stated above, please do not hesitate to contact me at 630-829-9623.

In accordance with Title 10 Code of Federal Regulations 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,


Frank P.D. Tran
License Reviewer
Materials Licensing Branch

Note: You could fax the response to 630-515-1078 or scan and email to frank.tran@nrc.gov.

License No. 13-06009-01

Docket No. 030-01625