

From: "Dimmick, Lisa" <lisa.dimmick@us.nucletron.com>
To: "Barbara Park" <bpark@mde.state.md.us>
Date: 3/21/2007 9:02:16 AM
Subject: RE: V3 SS&D

A physician can treat with fewer channels. So if the facility has a software license for 30 channels, they don't have to use all channels. They can still plan treatments with 18 or 24 channels for example.

Concerning the V3, yes only 3, 6, 30 channels will be offered as a configuration. Only customers who have the V2 (ie 18 channels) will be able to get their TCS with license for 18 channels (unless they want to upgrade to 30).

Regards,
Lisa

Lisa Dimmick
Director of QA& RA, RSO
Nucletron Corporation
Office: +1 443 545 2196
Mobile: +1 (443) 253 9172
lisa.dimmick@us.nucletron.com
Visit our website: www.nucletron.com

Nucletron supports the 12th International Conference "Optimal Use of Advanced Radiotherapy in Multimodality Oncology" in Rome. Please visit the website www.rome2007.org for further information.

This e-mail and any attachment sent with it are intended exclusively for the addressee(s), and may not be passed on to, or made available for use by any person other than the addressee(s). Nucletron Corporation rules out any and every liability resulting from any electronic transmission.

-----Original Message-----

From: Barbara Park [mailto:bpark@mde.state.md.us]
Sent: Wednesday, March 21, 2007 8:41 AM
To: Dimmick, Lisa
Subject: Re: V3 SS&D

I am re revising my description part of the SSD. So, according to the reply you sent, the V3 model 106.990 will be configurable to 3,6, or 30 channels only?

Only the 105.999 will be using 18 channels?

Would someone have to use all 30 channels or could they use less, like 18 or 24?

I just want to say the right thing about the 106.990.

Barbara