

**From:** [Lawyer, Dennis](#)  
**To:** [john.m.brisbin@dupont.com](mailto:john.m.brisbin@dupont.com)  
**Subject:** E.I. du Pont de Nemours and Company, Inc., Request for Additional Information Concerning Application for a License Amendment, Control 584013  
**Date:** Wednesday, July 02, 2014 8:29:00 AM

---

Dear Mr. John Brisbon,

This is in reference to your letter dated May 16, 2014, requesting for amendment to Nuclear Regulatory Commission License No. 07-13441-02, Docket No. 03020681. In order to continue our review, we need the following additional information:

1. Records from an inspection contained an inventory dated July 1, 1994, listing Room 117 for storage of carbon 14 material. Please perform a survey to show that this room can be released or state why this room does not need a survey.
2. In your "2013 Listing of Radioactive Material Use and Storage Areas at Glasgow Site," it shows room 213A as impacted. This room does not appear to have been performed in the DuPont Glasgow Facility Final Status Survey dated January 2014. Please perform the survey, submit missing information, or explain why this room does not additional surveying.
3. It appears that Room 182 is labeled as Room 184 in the DuPont Glasgow Facility Final Status Survey dated January 2014. Please confirm that Room 182 is labeled as Room 184 in the DuPont Glasgow Facility Final Status Survey or perform survey for Room 182 and submit.
4. Section 2.6 of DuPont Glasgow Facility Final Status Survey dated January 2014, states that a minimum of 12 sample points were measured in a rectangular grid. Based upon the drawing, It appears that: Room 182 only had 8 points in a rectangular grid; Room 184 only had 9 points in rectangular grid; and Rad Waste Storage Shed did not have a rectangular grid. Please perform the additional survey or discuss why the survey meets this criterion in section 2.6.
5. The DuPont Glasgow Facility Final Status Survey dated January 2014 shows scan surveys of building surfaces in impacted areas. The survey does not appear to show the percentage of building surfaces scanned. Please describe the percentage of area that was scanned in each room and the criteria being used for the amount scanned.

We will continue our review upon receipt of this information. Please reply to my attention at the Region 1 Office (Address below) and refer to Mail Control No. 584013. If you have technical questions regarding this letter, please call me at (610) 337-5366.

*Please note that you may not reply to this letter by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269). If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.*

Region 1 Office Mailing Address: Licensing Assistance Team, US Nuclear Regulatory

Commission Region I, 2100 Renaissance Boulevard, Suite 100, King of Prussia, PA  
19406-2713.

Dennis Lawyer  
Health Physicist  
U.S. Nuclear Regulatory Commission  
Division of Nuclear Material Safety  
610-337-5366  
610-337-5269 (F)