

From: <Jim.Hoey@vai-atl.ccmmail.valmet.com>
To: EPD-TP.Rad(elizabeth drinnon)
Date: 11/7/95 10:30pm
Subject: Prototype Testing

BWM

Elizabeth,

Under Prototype Testing, the temperature tests on the gauges were conducted according to ANSI 538 section 7.2.3:

The head was brought to the high temperature and held there for "stabilization time" plus one hour.

So the time period in the registry document should say "1 hour".

Regards,

Jim