

Endress+Hauser, Inc. Greenwood, Indiana USA

August 21, 2014

Lymari Sepulveda  
U.S. Nuclear Regulatory Commission  
FSME/MSSA/LB

**Amendment for Endress+Hauser SSDR NR-1302-D-101-B**

Dear Lymari,

Per your request dated August 19, 2014, I am submitting this as response to the requested information.

To question 1, "Please indicate if the prototype testing performed for the approved device is still applicable and that no detrimental effects will be caused as a result of the design change to the cover of the source capsules" - I confirm that the entire range of performed prototype testing is still applicable and that the changes do not affect any safety properties or ratings.

To question 2, "Please provide the dimensional units for the labels in drawing No. 960009440-A," -  
The dimensional units are all in millimeters.

Sincerely,  
Endress+Hauser, Inc.



Tim Thomas  
Level Product Business Manager