



CONVERSATION RECORD

07/29/2014

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

John W. May

DATE OF CONTACT

07/29/2014

TYPE OF CONVERSATION

☐ E-MAIL☒ TELEPHONE☐ INCOMING☒ OUTGOING

E-MAIL ADDRESS

~~mayj@wood.army.mil~~

john.w.may@us.army.mil

TELEPHONE NUMBER

(573) 563-6224

ORGANIZATION

Department of the Army

DOCKET NUMBER(S)

030-35257

LICENSE NUMBER(S)

24-32221-01

CONTROL NUMBER(S)

583967

SUBJECT

Request for additional information regarding letter dated May 19, 2014

SUMMARY

1) Mr. May stated that the licensee had only one sealed source model UDM-1A and the current activity is decayed from its original activity when the licensee received in the late '60.

2) Please provide a statement that before obtaining Troxler Model 3400 Series portable gauging devices, the Radiation Protection (Safety) Officer will have successfully completed a portable gauge training course provided by Troxler or equivalent course that meets Appendix D criteria in NUREG-1556, Volume 1, dated November 2001.

3) Please provide a statement that before using Troxler Model 3400 Series portable gauging devices, authorized users will have successfully completed a portable gauge training course provided by Troxler or equivalent course that meets Appendix D criteria in NUREG-1556, Volume 1, dated November 2001.

4) Please provide a statement that the licensee will implement and maintain procedures for routine maintenance of its gauges according to the manufacturer's recommendations and instructions.

Continue on Page 2

ACTION REQUIRED (IF ANY)

Please provide a response to Items 2 through 6 by August 8, 2014. The response must be dated and signed by an authorized representative. You could scan and e-mail the response to frank.tran@nrc.gov or fax it to 630-515-1078. To facilitate proper mail handling in our office, please refer Mail Control No. 583967 in your response.

If you have any questions regarding this request, please contact me at 630-829-9623 or e-mail to frank.tran@nrc.gov.

Continue on Page 3

NAME OF PERSON DOCUMENTING CONVERSATION

Frank Tran

SIGNATURE

CONVERSATION RECORD (continued)

SUMMARY: (Continued from page 1)

5) Please provide a statement that the licensee will send the gauge to the manufacturer or other person authorized by NRC or an Agreement State to perform non-routine maintenance or repair operations that require the removal of the source or source rod from the gauge.

6) Please provide a statement that the licensee will develop, implement, and maintain operating and emergency procedures that meet the criteria in the section entitled "Radiation Safety Program – Operating and Emergency Procedures" in NUREG-1556, Vol. 1, dated November 2001.

For your reference, NUREG-1556, Volume 1, dated November 2001 is published in the NRC public website at:
<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v1/r1/>