



CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to Infocollects.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

Shiawassee County Road Commission
701 W. Corunna Avenue
Corunna, MI 48817

LICENSE NUMBER

21-26008-01

DOCKET NUMBER

030-30972

LICENSE EXPIRATION DATE

January 31, 2017

A. LICENSE STATUS (Check the appropriate box)

- ☐ This license has expired. ☒ This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- ☐ 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- ☒ 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
- ☒ a. Transfer of radioactive materials to the licensee listed below:
The transfer occurred on March 12, 2015, to Spicer Group, Inc of NRC License No. 21-26099-01, the contact for this license is Rick E. Born, (989) 754-4717.
- ☐ b. Disposal of radioactive materials:
- ☐ 1. Directly by the licensee:
- ☐ 2. By licensed disposal site:
- ☐ 3. By waste contractor:
- ☐ c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- ☐ 1. A radiation survey was conducted by the licensee. The survey confirms:
- ☐ a. the absence of licensed radioactive materials
- ☐ b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- ☐ 2. A copy of the radiation survey results:
- ☐ a. is attached; or ☐ b. is not attached (Provide explanation); or ☐ c. was forwarded to NRC on: _____ Date _____
- ☒ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- ☒ a. The results of the latest leak test are attached; and/or ☒ b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Brent Friess	Superintendent/Interim Managing Director	989-743-2228	bfriess@shiawasseeeroads.com

Mail all future correspondence regarding this license to:

701 W. Corunna Avenue, Corunna, MI 48817

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE

Brent Friess, Superintendent/Interim Managing Director

SIGNATURE

Brent P. Friess

DATE

3/16/15

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

InstroTek, Inc.
5908 Triangle Drive
Raleigh, NC 27617
(919)875-8371 Fax (919)875-8328

2/18/2015
Test Number: 1

SARAH WHITING
Rowe Professional Services Company
540 S. Saginaw Street SUITE 200
P.O. BOX3748
Flint, MI 48502

Phone: (810)341-7500
Fax: (810)341-7573

LEAK TEST CERTIFICATE

NC Materials License #092-1073-1


This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3440
Gauge S/N: 31945


Test Date: 2/11/2015

Source (Model/Serial#)	Reading in microCuries
47-3358(40)	0.00000
750-7002(8)	0.00022

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source.* The source(s) tested above may remain in use.

Reviewed by: 

Date: 02/18/2015

Customer Signature: 

Date: 2/25/15

*CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 3 mCi Cs-137.

SHIAWASSEE COUNTY ROAD COMMISSION
701 W CORUNNA AVENUE
CORUNNA, MI 48817-1229

GRAND RAPIDS MI 494

17 MAR 2015 PM 2 L



Materials Licensing Section
U. S. Nuclear Regulatory Commission, Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

60532435260

