

SEALED SOURCE LEAK TEST AND INVENTORY

Licensee: Cardiovascular Associates of the Virgin Islands, PLLC
 Address: 4100 Sion Farm Shopping Center, Suite 5
 City: Christiansted State: St. Croix, V.I. Zip: 00820
 License #: 55-31325-01 Telephone: (340) 713-7000

Sample Identification/Analysis

Source

Nuclide	Co-57	Co-57	Cs-137
Mfg	IPL	IPL	IPL
Model	NES 392	RV-057-5M	RV-137-200U
Serial	1455-016	1461-35-9	1452-43-11
Assay Date	04/01/2011	03/01/2011	10/01/2010
Activity	10 mCi	5.727 mCi	197.7 uCi

Current

Activity	2.3 mCi	1.22 mCi	188.5 uCi
Location	Hot Lab	Hot Lab	Hot Lab
Use	Camera	Dose Calibrator	Dose Calibrator

Wipe

Date	10/29/2012	10/29/2012	10/29/2012
By	Jason Seav	Jason Seav	Jason Seav

Assay

Date	11/07/2012	11/07/2012	11/07/2012
μCi	< 0.0001	< 0.0001	< 0.0001

LLD

μCi	0 net cpm	0 net cpm	0 net cpm
-----	-----------	-----------	-----------

Analysis

<0.005 μCi	x	x	x
>0.005 μCi			

Next Leak Test Due: (every 6 mo.) 04/2013
 Next Inventory Due: (every 3 mo.) 01/2013

Comments/Actions:

Assay Performed By: Microtec Services, Inc. Date: 11/07/2012

License # 104636 Agency: Texas

Assay Approved By: Charles H. Rose Date: 11/07/2012

Charles H. Rose, MA, MSPH, D(ABSNM)

RSO Signature: _____ Date: _____



Nuclear Cardiology Systems, Inc. (NC Systems)

5660 Airport Blvd., Suite 101, Boulder, CO 80301 • (800) 548-4024 • (303) 541-0044 • (303) 541-0066 FAX