



Region I Office
Division of Nuclear Materials Safety
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713
(610)337-5000

Telephone Conversation Record

Date: April 15, 2015

License No. 06-27909-01

Docket No.(no hyphens): 03029568

Mail Control/Report No. 586316

Licensee Name: Central Connecticut Cardiologists, LLC

Participant(s) Name/Title: Peter Mas, Consultant

Work Telephone No. 860-972-2676

Business Cellphone No.

NRC Representative Name/Title: Farrah Gaskins, Health Physicist

Subject: Clarification of Additional Information

Discussion: On 4/10/15, I contacted Peter Mas to ask for leak tests of sealed sources that were being held by the pharmacy and I also asked for calibration dates of the survey instruments used to perform surveys in the areas they wished to release for unrestricted use. There was discrepancy on the calibration dates in that both the calibration date and the calibration due date were 8/25/2015. I contacted Peter who confirmed that the calibration date SHOULD BE 8/25/2014 and the calibration due date SHOULD BE 8/25/2015

Action Required: Place into Adams.

SUNSI REVIEW

Document Availability: ☒ Public or ☐ Non-Public

Document Sensitivity:

(select "1" value to the right)

☒ Non-Sensitive

☐ Non-Sensitive Copyright

☐ Sensitive – Proprietary

☐ Sensitive – Privacy Act (includes Personally Identifiable Information (PII))

☐ Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.)

☐ Sensitive – Security-Related-Periodic Review Required

☐ MD 3.4 Non-Public B.1 (Non-Sensitive)

☐ MD 3.4 Non-Public A.3 (Sensitive Security Related)

☐ MD 3.4 Non-Public A.4 (Sensitive Proprietary)

☐ MD 3.4 Non-Public A.5 (Sensitive-Privacy Act)

☐ MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc)

☐ MD 3.4 Non-Public A.7 (Internal)

SUNSI Review Completed by: FGaskins

: If applicable, add OFFICIAL USE ONLY – SECURITY RELATED INFORMATION banners and change SUNSI review to A.3.