

## SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

## 1. LICENSEE/LOCATION INSPECTED:

Mirview, Inc.  
3883 Ethan Lane  
Orlando, FL 32814  
Location Inspected: Eaton Rapids, MI

REPORT NUMBER(S) 2015-001

## 2. NRC/REGIONAL OFFICE

Region III  
U. S. Nuclear Regulatory Commission  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

## 3. DOCKET NUMBER(S)

030-38736

## 4. LICENSE NUMBER(S)

09-26775-01

## 5. DATE(S) OF INSPECTION

April 24, 2015

## LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

## Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Ryan Craffey		4/27/15
BRANCH CHIEF	Aaron McCraw		5/6/15

**Docket File Information**

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:

Mirview, Inc.  
3883 Ethan Lane  
Orlando, FL 32814  
Location Inspected: Eaton Rapids, MI

REPORT NUMBER(S) 2015-001

2. NRC/REGIONAL OFFICE

Region III  
U. S. Nuclear Regulatory Commission  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-38736

4. LICENSE NUMBER(S)

09-26775-01

5. DATE(S) OF INSPECTION

April 24, 2015

6. INSPECTION PROCEDURES USED

87130

7. INSPECTION FOCUS AREAS

All

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)

02220

2. PRIORITY

3

3. LICENSEE CONTACT

Ray Carlson, MS - RSO

4. TELEPHONE NUMBER

(734) 455-4730

☐

Main Office Inspection

Next Inspection Date: 04/24/2018

☐

Field Office Inspection

☒

Temporary Job Site Inspection 1500 South Main Street, Eaton Rapids, MI

**PROGRAM SCOPE**

This was an unannounced routine inspection of a mobile nuclear medicine service provider authorized to conduct diagnostic administrations of radiopharmaceuticals at temporary job sites within NRC jurisdiction, and to store sealed sources for instrument calibration at the RSO's office in Ann Arbor, Michigan. At the time of the inspection, one nuclear medicine technologist performed 1-2 administrations per day beginning at 7:00 am every Monday, Friday, and alternating Wednesdays (including April 15, 2015) on the premises of Eaton Rapids Medical Center (ERMC) in Eaton Rapids, Michigan. The technologist indicated that the licensee did not currently operate at any other job sites, but that it was in talks with ERMC to expand its operations there to Mondays through Fridays. A local radiopharmacy delivered unit doses directly to the licensee's trailer, where the technologist was required to be present in order to receive them. The licensee's medical physicists conducted quarterly audits of the radiation safety program.

**PERFORMANCE OBSERVATIONS**

The inspector toured the nuclear medicine trailer to evaluate the licensee's measures for materials security, hazard communication and exposure control. The inspector conducted independent and confirmatory surveys in and around the trailer, and found no evidence of residual contamination or exposures to members of the public in excess of regulatory limits. The inspector observed the preparation and administration of one hepatobiliary scan using technetium-99m. The technologist also demonstrated and discussed the implementation of licensee procedures for package receipt, waste handling, area surveys and spill response, as well as cardiac stress tests within the ERMC facility. Through these and other discussions, the inspector found the technologist to be knowledgeable of radiation protection principles, regulatory requirements and licensee procedures.

The inspector also reviewed a selection of licensee records, including quarterly audits, dosimetry, training certificates, its service agreement with ERMC, as well as documentation of instrument calibration, package receipt and area surveys.

No violations of NRC requirements were identified as a result of this inspection.