



Order No. EA-12-051

RS-15-120

June 9, 2015

U.S. Nuclear Regulatory Commission
ATTN: Document Control Desk
Washington, DC 20555-0001

Braidwood Station, Units 1 and 2
Facility Operating License Nos. NPF-72 and NPF-77
NRC Docket Nos. STN 50-456 and STN 50-457

Subject: Report of Full Compliance with March 12, 2012 Commission Order Modifying Licenses with Regard to Reliable Spent Fuel Pool Instrumentation (Order Number EA-12-051)

References:

1. NRC Order Number EA-12-051, "Issuance of Order to Modify Licenses with Regard to Reliable Spent Fuel Pool Instrumentation," dated March 12, 2012
2. NRC Interim Staff Guidance JLD-ISG-2012-03, "Compliance with Order EA-12-051, Reliable Spent Fuel Pool Instrumentation," Revision 0, dated August 29, 2012
3. NEI 12-02, Industry Guidance for Compliance with NRC Order EA-12-051, "To Modify Licenses with Regard to Reliable Spent Fuel Pool Instrumentation," Revision 1, dated August 2012
4. Exelon Generation Company, LLC's Initial Status Report in Response to March 12, 2012 Commission Order Modifying Licenses with Regard to Requirements for Reliable Spent Fuel Pool Instrumentation (Order Number EA-12-051), dated October 25, 2012 (RS-12-181)
5. Exelon Generation Company, LLC Overall Integrated Plan in Response to March 12, 2012 Commission Order Modifying Licenses with Regard to Reliable Spent Fuel Pool Instrumentation (Order Number EA-12-051), dated February 28, 2013 (RS-13-027)
6. NRC letter to Exelon Generation Company, LLC, Request for Additional Information Regarding Overall Integrated Plan for Reliable Spent Fuel Pool Instrumentation, dated July 11, 2013
7. Exelon Generation Company, LLC letter to NRC, Response to Request For Additional Information - Overall Integrated Plan in Response to Commission Order Modifying License Requirements for Reliable Spent Fuel Pool Instrumentation (Order EA-12-051), dated July 31, 2013 (RS-13-194)
8. Exelon Generation Company, LLC First Six-Month Status Report in Response to March 12, 2012 Commission Order Modifying Licenses with Regard to Reliable Spent Fuel Pool Instrumentation (Order Number EA-12-051), dated August 28, 2013 (RS-13-112)

9. Exelon Generation Company, LLC Second Six-Month Status Report in Response to March 12, 2012 Commission Order Modifying Licenses with Regard to Reliable Spent Fuel Pool Instrumentation (Order Number EA-12-051), dated February 28, 2014 (RS-14-017)
10. Exelon Generation Company, LLC Third Six-Month Status Report in Response to March 12, 2012 Commission Order Modifying Licenses with Regard to Reliable Spent Fuel Pool Instrumentation (Order Number EA-12-051), dated August 28, 2014 (RS-14-195)
11. Exelon Generation Company, LLC Fourth Six-Month Status Report in Response to March 12, 2012 Commission Order Modifying Licenses with Regard to Reliable Spent Fuel Pool Instrumentation (Order Number EA-12-051), dated February 27, 2015 (RS-15-026)
12. NRC letter to Exelon Generation Company, LLC, Braidwood Station, Units 1 and 2 – Interim Staff Evaluation and Request for Additional Information Regarding the Overall Integrated Plan for Implementation of Order EA-12-051, Reliable Spent Fuel Pool Instrumentation (TAC Nos. MF0821 and MF0822), dated November 4, 2013
13. NRC letter to Exelon Generation Company, LLC, Braidwood Station, Units 1 and 2 – Report for the Audit Regarding Implementation of Mitigating Strategies and Reliable Spent Fuel Pool Instrumentation Related to Orders EA-12-049 and EA-12-051 (TAC Nos. MF0895 and MF0896), dated May 27, 2015

On March 12, 2012, the Nuclear Regulatory Commission (“NRC” or “Commission”) issued Order EA-12-051, “Order Modifying Licenses with Regard to Reliable Spent Fuel Pool Instrumentation,” (Reference 1) to Exelon Generation Company, LLC (EGC). Reference 1 was immediately effective and directed EGC to install reliable spent fuel pool level instrumentation. Specific requirements are outlined in Attachment 2 of Reference 1.

Reference 1 required submission of an initial status report 60 days following issuance of the final interim staff guidance (Reference 2) and an overall integrated plan (OIP) pursuant to Section IV, Condition C. Reference 2 endorsed industry guidance document NEI 12-02, Revision 1 (Reference 3) with clarifications and exceptions identified in Reference 2. Reference 4 provided the EGC initial status report regarding reliable spent fuel pool instrumentation. Reference 5 provided the Braidwood Station, Units 1 and 2 OIP.

Reference 1 required submission of a status report at six-month intervals following submittal of the OIP. References 8, 9, 10, and 11 provided the first, second, third, and fourth six-month status reports, respectively, pursuant to Section IV, Condition C.2, of Reference 1 for Braidwood Station.

The purpose of this letter is to provide the report of full compliance with the March 12, 2012 Commission Order Modifying Licenses with Regard to Reliable Spent Fuel Pool Instrumentation (Order Number EA-12-051) (Reference 1) pursuant to Section IV, Condition C.3 of the Order for Braidwood Station, Units 1 and 2.

Braidwood Station has installed two independent full scale level monitors for the Spent Fuel Pool (SFP) in response to Order EA-12-051. Braidwood Station OIP Open Items have been addressed and closed as documented in References 8, 9, 10, and 11, and are considered complete pending NRC Closure. The information provided herein documents full compliance for Braidwood Station, Units 1 and 2 with Reference 1.

EGC's response to the NRC OIP Requests for Additional Information (OIP RAIs), and the NRC Interim Staff Evaluation (ISE) Open Items (ISE RAIs) identified in References 6 and 12 have been addressed and closed as documented in References 7, 8, 9, 10, and 11, and below, and are considered complete pending NRC Closure. The following table provides completion references for each NRC OIP RAI and ISE RAI.

OIP Open Item No. 2	Reference 8
OIP RAI Nos. 1a, 2, 3, 4, 5, 6, 7b, 8a, 8b, 8c, 10, 11a	Reference 10
OIP RAI Nos. 7a, 11b, 11c	References 10 and 11
OIP RAI No. 1b	Reference 7
OIP RAI Nos. 8d, 8e	With this submittal as provided below
OIP RAI No. 9	Reference 8
ISE RAI Nos. 1, 3, 5, 6, 13	Reference 10
ISE RAI No. 8	References 10 and 11

Note: ISE RAIs are not duplicated in the table above if previously issued as OIP RAIs in Reference 6.

EGC's response to the NRC audit questions and any additional audit open items have been addressed and closed as documented in the NRC Site Audit Report (Reference 13). Reference 13 contains no remaining audit open items regarding Braidwood Station's compliance with NRC Order EA-12-051.

The table below documents the completion of the final remaining open actions as identified in Reference 11. As stated above, EGC provides the response for the following items and considers them to be complete for Braidwood Station.

Item	Description	Reference
OIP Item 10 (RAI-8, Ref. 6)		
d) A discussion as to how these surveillances will be incorporated into the plant surveillance program.	Braidwood Station has developed calibration, functional test, and channel verification procedures for the SFPI per Westinghouse recommendations to ensure reliable, accurate, and continuous SFPI functionality.	<u>Complete</u>
e) A description of the preventive maintenance tasks required to be performed during normal operation, and the planned maximum surveillance interval that is necessary to assure	Braidwood Station has developed preventive maintenance tasks for the SFPI per Westinghouse recommendation identified in the technical manual WNA-GO-00127-GEN (Attachment 2, item 19) to assure that the channels are fully conditioned to	<u>Complete</u>

that the channels are fully conditioned to accurately and reliably perform their functions when needed.	accurately and reliably perform their functions when needed.	
---	--	--

MILESTONE SCHEDULE – ITEMS COMPLETE

Milestone	Completion Date
Submit 60 Day Status Report	October 25, 2012
Submit Overall Integrated Plan	February 28, 2013
Submit Responses to RAIs	July 31, 2013
Submit 6 Month Updates:	
Update 1	August 28, 2013
Update 2	February 28, 2014
Update 3	August 28, 2014
Provide Final Safety Evaluation (SE) Information	September 30, 2014
Update 4	February 27, 2015
Modifications:	
Conceptual Design	3Q2012
Issue Exelon Fleet contract to procure SFPI Equipment	2Q2013
Begin Detailed Design Engineering	4Q2013
Complete and Issue SFPI Modification Package	3Q2014
Begin Installation	1Q2015
Complete SFPI Installation and Put Into Service	April 28, 2015

ORDER EA-12-051 COMPLIANCE ELEMENTS SUMMARY

The elements identified below for Braidwood Station, as well as the site overall integrated plan response submittal (Reference 5), the 6-Month Status Reports (References 8, 9, 10, and 11), and any additional docketed correspondence, demonstrate compliance with Order EA-12-051.

IDENTIFICATION OF LEVELS OF REQUIRED MONITORING - COMPLETE

Braidwood Station has identified the three required levels for monitoring SFP level in compliance with Order EA-12-051. These levels have been integrated into the site processes for monitoring level during events and responding to loss of SFP inventory.

INSTRUMENT DESIGN FEATURES - COMPLETE

The design of the instruments installed at Braidwood Station complies with the requirements specified in the Order and described in NEI 12-02 "Industry Guidance for Compliance with NRC Order EA-12-051". The instruments have been installed in accordance with the station design control process.

The instruments have been arranged to provide reasonable protection against missiles. The instruments have been mounted to retain design configuration during and following the maximum expected ground motion. The instruments will be reliable during expected environmental and radiological conditions when the SFP is at saturation for extended periods. The instruments are independent of each other and have separate and diverse power supplies. The instruments will maintain their design accuracy following a power interruption and are designed to allow for routine testing and calibration.

The instrument display is readily accessible during postulated events and allows for SFP level information to be promptly available to decision makers.

PROGRAM FEATURES - COMPLETE

Training for Braidwood Station, Units 1 and 2 has been completed in accordance with an accepted training process as recommended in NEI 12-02, Section 4.1.


Operating and maintenance procedures for Braidwood Station have been developed and integrated with existing procedures. Procedures have been verified and are available for use in accordance with the site procedure control program.

Site processes have been established to ensure the instruments are maintained at their design accuracy.

This letter contains no new regulatory commitments. If you have any questions regarding this report, please contact David P. Helker at 610-765-5525.

I declare under penalty of perjury that the foregoing is true and correct. Executed on the 9th day of June 2015.

Respectfully submitted,



Glen T. Kaegi
Director - Licensing & Regulatory Affairs
Exelon Generation Company, LLC