



**UNITED STATES  
NUCLEAR REGULATORY COMMISSION**  
REGION I  
2100 RENAISSANCE BOULEVARD, SUITE 100  
KING OF PRUSSIA, PA 19406-2713

June 29, 2015

Docket No. 03003289  
EA-15-131

License No. 44-10187-03

Christina Oliver  
Vice President of Clinical Services  
University of Vermont Medical Center  
111 Colchester Avenue  
Burlington, VT 05401

**SUBJECT: NRC INSPECTION REPORT NO. 03003289/2015001, UNIVERSITY OF VERMONT MEDICAL CENTER, BURLINGTON, VERMONT SITE, NOTICE OF VIOLATION, AND EXERCISE OF ENFORCEMENT DISCRETION**

Dear Ms. Oliver:

On June 2-3, and 22, 2015, Robert L. Gallagher of this office conducted a safety inspection at the above address of activities authorized by the above listed NRC license. The inspection was an examination of your licensed activities as they relate to radiation safety and to compliance with the Commission's regulations and the license conditions. The inspection consisted of observations by the inspector, interviews with personnel, and a selective examination of representative records. The findings of the inspection were discussed with you at the conclusion of the inspection.

Based on the results of this inspection and in accordance with the NRC Enforcement Policy, the NRC has determined that two violations of NRC requirements occurred. The first violation involved 10 CFR 35.60 which requires, in part, that licensees calibrate the instrument used to measure the activity of the dosage before it is administered to each patient or human research subject. This calibration may either be performed in accordance with nationally recognized standards or calibration instructions provided by the manufacturer. Contrary to the above, the University of Vermont Medical Center (UVMC) was unable to calibrate a Rubidium-82 (Rb-82) generator unit in accordance with the regulations because there are currently neither nationally recognized standards nor specific calibration procedures for calibrating detectors in a dynamic mode (i.e., while liquids are flowing past the detector). Until such standards or procedures are developed, compliance with 10 CFR 35.60 is not possible. The second violation involved 10 CFR 35.63 which requires, in part, that a licensee determine the activity of each dosage administered before medical use. Due to the 76 second half-life of Rb-82 and direct infusion of the Rb-82 into the patient, users of this generator system are unable to measure patient dosage of Rb-82 prior to administration.

Although violations of 10 CFR 35.60 and 10 CFR 35.63 were identified, which, in accordance with the NRC Enforcement Policy, would normally be categorized at Severity Level IV, UVMC met all of the criteria in NRC Enforcement Guidance Memorandum (EGM) 13-003, "Interim Guidance for Dispositioning Violations Involving 10 CFR 35.60 and 10 CFR 35.63 for the Calibration of Instrumentation to Measure the Activity of Rubidium-82 and the Determination of Rubidium-82 Patient Doses," dated April 18, 2013. Therefore, the NRC is exercising

enforcement discretion and not pursuing any enforcement action for these violations. The NRC's decision is based on the criteria listed in EGM 13-003. Specifically, (1) UVMC has written test procedures to ensure that the infusion pump flow rate is consistent and accurate, and that the radiation detector meets the manufacturer's specifications; (2) UVMC has confirmed that the required infusion rate and radiation detector tests were performed within the last twelve months, and has maintained records documenting the performance and results of these tests; (3) all authorized users for medical uses under 10 CFR 35.200 who are using the Rb-82 generator infusion cart, as well as the Radiation Safety Officer, have successfully completed training specific to the manufacturer and model of the generator and infusion cart being used, and documentation of satisfactory completion of such training has been maintained; and (4) UVMC has recorded the activity of each dosage administered, as provided by the infusion cart.

On June 22, 2015, Mr. Gallagher informed your staff of this conclusion by telephone and stated that a letter documenting the conclusion would be issued. No further action or response is required on your part. In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and your response, if you choose to provide one, will be made available electronically for public inspection in the NRC Public Document Room or from the NRC document system (ADAMS), accessible from the NRC website at <http://www.nrc.gov/reading-rm/adams.html>. To the extent possible, your response should not include any personal privacy, proprietary, or safeguards information so that it can be made available to the Public without redaction.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Regulations, Guidance and Communications**. The current Enforcement Policy is included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **About NRC, Organizations & Functions; Office of Enforcement; Enforcement documents**; then **Enforcement Policy (Under 'Related Information')**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

Please contact Mr. Gallagher at (610) 337-5182 if you have any questions regarding this matter.

Sincerely,

*/RA J. Nick for/*

Daniel S. Collins, Director  
Division of Nuclear Materials Safety

cc: Marleen M. Moore, M.S.,  
Radiation Safety Officer  
State of Vermont

enforcement discretion and not pursuing any enforcement action for these violations. The NRC's decision is based on the criteria listed in EGM 13-003. Specifically, (1) UVMC has written test procedures to ensure that the infusion pump flow rate is consistent and accurate, and that the radiation detector meets the manufacturer's specifications; (2) UVMC has confirmed that the required infusion rate and radiation detector tests were performed within the last twelve months, and has maintained records documenting the performance and results of these tests; (3) all authorized users for medical uses under 10 CFR 35.200 who are using the Rb-82 generator infusion cart, as well as the Radiation Safety Officer, have successfully completed training specific to the manufacturer and model of the generator and infusion cart being used, and documentation of satisfactory completion of such training has been maintained; and (4) UVMC has recorded the activity of each dosage administered, as provided by the infusion cart.

On June 22, 2015, Mr. Gallagher informed your staff of this conclusion by telephone and stated that a letter documenting the conclusion would be issued. No further action or response is required on your part. In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and your response, if you choose to provide one, will be made available electronically for public inspection in the NRC Public Document Room or from the NRC document system (ADAMS), accessible from the NRC website at <http://www.nrc.gov/reading-rm/adams.html>. To the extent possible, your response should not include any personal privacy, proprietary, or safeguards information so that it can be made available to the Public without redaction.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Regulations, Guidance and Communications**. The current Enforcement Policy is included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **About NRC, Organizations & Functions; Office of Enforcement; Enforcement documents**; then **Enforcement Policy (Under 'Related Information')**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

Please contact Mr. Gallagher at (610) 337-5182 if you have any questions regarding this matter.

Sincerely,

**/RA J. Nick for/**

Daniel S. Collins, Director  
Division of Nuclear Materials Safety

cc: Marleen M. Moore, M.S.,  
Radiation Safety Officer  
State of Vermont

**DISTRIBUTION:**

OEMail

NHasan, OE

LSreenivas, OE

MBurgess, FSME

RSun, FSME

BBickett, RI

MMcLaughlin, RI

CCrisden, RI

DOCUMENT NAME: G:\WordDocs\Current\Insp Letter\L44-10187-03.2015001.docx

**SUNSI Review Complete:** RGallagher

ML15182A403

After declaring this document An Official Agency Record it **will** be released to the Public.

To receive a copy of this document, indicate in the box: **C** = Copy w/o attach/encl **E** = Copy w/ attach/encl **N** = No copy

OFFICE	DNMS/RI	N	DNMS/RI		ORA/RI		DNMS/RI	
NAME	RGallagher/rlg		JDwyer/jpd		BBickett/mm f/		DCollins/jln	
DATE	06/26/15		06/26/15		06/26/15		06/29/15	

OFFICIAL RECORD COPY