

License: 49-10982-02
Docket: 030-13772
Control: 556 581

Cook, Jackie

From: Cook, Jackie
Sent: Monday, July 13, 2015 2:26 PM
To: Tom Nance (tomnance@sheridanhospital.org)
Cc: Katanic, Janine; Dykert, Jason; Torres, Roberto J
Subject: Amendment Request dated April 14, 2015 & former license condition 14

Importance: High

Good Afternoon Tom: I apologize for not letting you know why former license condition 14 was removed on Amendment No. 30 Corrected Copy dated April 24, 2014. Please note that former license condition 14 was removed from your license because it appears that any sealed sources authorized under your license are associated with 10 CFR Part 35 use (medical) and the leak test requirements for these sealed sources are found in 10 CFR 35.67, Requirements for possession of sealed sources and brachytherapy sources. Please note that the Co-57 flood sources and any other calibration and reference sources are authorized under 10 CFR 35.65 (which will make them for "medical use" also).

Please note in amendment request dated April 14, 2015, you requested former license condition 14 to be restored to your license. However, based on our earlier telephone conversation and the information noted above, we will void your amendment request because it is not necessary.

If our understanding is not correct and you do have sealed sources authorized under your license not associated with medical use, please contact our office immediately and we will issue an amendment and/or corrected copy accordingly.

Please do not hesitate to contact me at your convenience at the phone numbers and/or email address listed below if I can be of further assistance or you have any more questions.

Warmest Regards,

Ms. Jacqueline "Jackie" D. Cook
Senior Health Physicist
US Nuclear Regulatory Commission Region IV
Division of Nuclear Materials Safety
Nuclear Materials Safety Branch B
1600 East Lamar Blvd., Arlington, TX 76011
817-200-1132 (office)/817-200-1263 (fax)
Email address: Jackie.Cook@nrc.gov

PUBLIC

- ☐ Immediate Release
☒ Normal Release

NON-PUBLIC

- ☐ A.3 Sensitive-Security Related
☐ A.7 Sensitive Internal
☐ Other:

Reviewer: JAC Date: 7/13/15