

From: [Weidner, Tara](#)
To: jseibyl@mnimaging.com
Subject: MNI - NRC amendment
Date: Wednesday, July 15, 2015 2:47:00 PM

License No.: 06-30624-01
Docket No.: 03035657
Control No.: 586920

***** UPON RECEIPT, PLEASE CONFIRM THE RECEIPT OF
THIS E-MAIL *****

Good afternoon Dr. Seibyl.

I am reviewing your request for an additional PET/CT camera and renovation of the 8th floor. I was confused by the diagrams and information provided. However, I spoke to Mike Bohan and he cleared up my confusion. Based on the information that Mike gave me, I understand that you are renovating the area where the PET and SPECT cameras are currently located. You are moving the PET camera to former office space adjacent to its current location. In the area formerly occupied by the PET and SPECT cameras, you are installing a new PET/CT camera. Where will the SPET camera be located? No changes are being made to the areas where radioactive material is prepared and stored (ie. the existing hot lab). In addition, I have reviewed the shielding calculations performed by Mike and the recommended amount of shielding. In a letter, signed by you, please indicate whether or not my understanding of the renovations are accurate and confirm that the recommended amount of shielding will be installed during construction. Please provide this information by July 27, 2015. You may scan the letter and forward it to my e-mail.

Thanks in advance,
Tara L. Weidner
Senior Health Physicist, US NRC