

**From:** [Nguyen, Janice](#)  
**To:** ["Barreca, Leandro J."](#)  
**Subject:** RE: RE: RE: RE: RE: RE: NRC Request for Additional Information - West Virginia Radiation Therapy Services, Inc.  
- Mail Control Number 586882  
**Date:** Friday, July 24, 2015 11:51:50 AM

---

Licensee: West Virginia Radiation Therapy Services, Inc.  
License No.: 47-25296-01  
Docket No.: 030-33538  
Mail Control No.: 586882

Hi Leandro,

Thanks so much for the response.

I do have additional questions.

- 1) Is the area adjacent to the treatment room located outside (i.e., parking lot)?
- 2) In Item 5, it was stated that required surveys will be performed with the Ludlum 14C survey instrument with pancake probe. What type of instrumentation will be used for contamination surveys (to count package wipes and weekly wipes)? If the GM will be used to count wipes, please confirm that measurements will be conducted in a low background area. Please also provide the minimum detectable activity (MDA) for the meter. Please note that DOT's revised required MDA is 24.

If you could please provide a short, signed response, it would be greatly appreciated.

Thank you!

Jan Nguyen

Janice Nguyen  
Senior Health Physicist  
Division of Nuclear Materials Safety  
U.S. Nuclear Regulatory Commission Region I  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406-2713  
(610) 337-5006  
FAX (610) 337-5269