

July 28, 2015

Janice Nguyen  
Senior Health Physicist  
Division of Nuclear Material Safety  
U.S. Nuclear Regulatory Commission, Region I  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406-2713

RE: Amendment Request to Radioactive Materials License # 47-25296-01  
Mail Control No. 586882

Dear Ms. Nguyen:

This is in reference to your email dated July 24, 2015, requesting to more information to amend Nuclear Regulatory Commission License No. 47-25296-01 in order to change our Radiation Safety Officer, to add the use of Radium 223 permitted by 10 CFR 35.300, and to remove the high dose rate remote afterloader unit.

1. The area adjacent to the treatment room located outside is a parking lot area.
2. We confirm that the GM will be used to count wipes. We confirm that measurements will be conducted in a low background area providing that the minimum detectable activity (MDA) for the meter is 24 as required by DOT's revised required MDA.
3. The email address for Dr. Baisden is [jbaisden@rtsx.com](mailto:jbaisden@rtsx.com)

Should you have any questions or comments, feel free to contact me at (239) 768-7377.

Sincerely,

A handwritten signature in black ink, appearing to read 'Leandro J. Barreca', with a stylized flourish at the end.

Leandro J. Barreca, MSEM  
Physics Department