



UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

TELEFAX TRANSMITTAL

DATE July 15, 2015

NUMBER OF PAGES 2

SEND TO Robert Dahman, Radiation Safety Officer, Indiana Department of
Transportation, NRC License 13-26345-01

LOCATION Fort Wayne, Indiana

FAX NUMBER (260) 969-8279

☐ VERIFY BY CALLING

FROM: Bill Reichhold
(Sender)

TELEPHONE NUMBER (630) 829-9839

FAX NUMBER (630) 515-1078

If you do not receive the complete fax transmittal, please contact the sender as
soon as possible at the telephone number provided above.

MESSAGE See accompanying documents.

NOTICE

This message is intended only for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, or exempt from disclosure under applicable law. If the reader of this message is not the intended recipient or the employee responsible for delivering the message to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. If you received this communication in error, please notify the sender immediately by telephone and return the original to the above address, by U.S. Mail. Thank You.

The following additional information is needed to complete our review of your request to terminate your NRC license.

1. Please submit a history of all radionuclides used at your facility. For example, we only used cesium-137 and americium-241 as sealed sources in portable nuclear gauges. Our NRC license authorized up to 56 Troxler moisture/density gauges and 2 Humboldt Scientific Model 5001 gauges, however, we only receive and possessed 18 Troxler moisture/density gauges. Our NRC license authorized californium-252, however, we never received or possessed any californium-252.
2. Please submit a copy of the leak tests for the 18 moisture/density gauges that was current when the gauges were transferred to InstroTek, Inc. on August 7, 2013. Please specify if you had any history of leaking sealed sources. If you never had any leaking sealed sources, please state so.
3. Please confirm all radionuclides have been removed from your former storage locations and transferred to InstroTek, Inc.
4. Please specify the contact and the mailing address where you wish the terminated NRC license sent to. If you wish the terminated NRC license sent to the address on the current license, please state so.

Please send a facsimile (630-515-1078) of your response to the above within 14 days and state, Response to Control 587056. Please include a cover letter on company letterhead, dated and signed (signed by an individual who is authorized to sign official documents on behalf of the licensee) with your response letter. Please call me at 630-829-9839 if you have any questions.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this facsimile and the attached documents will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

From the desk of:



Bill Reichhold