

NRC FORM 313

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0120

EXPIRES: 05/31/2015

(03-2014)  
10 CFR 30, 32, 33, 34  
35, 36, 37, 39, and 40APPLICATION FOR MATERIALS  
LICENSE

Estimated burden per response to comply with this mandatory collection request: 4.3 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (1-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to [Infocollections.Resource@nrc.gov](mailto:Infocollections.Resource@nrc.gov), and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION. SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW. \*AMENDMENTS/RENEWALS THAT INCREASE THE SCOPE OF THE EXISTING LICENSE TO A NEW OR HIGHER FEE CATEGORY WILL REQUIRE A FEE.

## APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH:

OFFICE OF FEDERAL & STATE MATERIALS AND  
ENVIRONMENTAL MANAGEMENT PROGRAMS  
DIVISION OF MATERIALS SAFETY AND STATE AGREEMENTS  
U.S. NUCLEAR REGULATORY COMMISSION  
WASHINGTON, DC 20555-0001

## ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS:

## IF YOU ARE LOCATED IN:

ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA,  
KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY,  
NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH  
CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA,

## SEND APPLICATIONS TO:

LICENSING ASSISTANCE TEAM  
DIVISION OF NUCLEAR MATERIALS SAFETY  
U.S. NUCLEAR REGULATORY COMMISSION, REGION I  
2100 RENAISSANCE BOULEVARD, SUITE 100  
KING OF PRUSSIA, PA 19406-2713

## IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN,  
SEND APPLICATIONS TO:

MATERIALS LICENSING BRANCH  
U.S. NUCLEAR REGULATORY COMMISSION, REGION III  
2443 WARRENVILLE ROAD, SUITE 210  
LISLE, IL 60532-4352

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS,  
LOUISIANA, MISSISSIPPI, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH  
DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS,  
UTAH, WASHINGTON, OR WYOMING,

## SEND APPLICATIONS TO:

NUCLEAR MATERIALS LICENSING BRANCH  
U.S. NUCLEAR REGULATORY COMMISSION, REGION IV  
1600 E. LAMAR BOULEVARD  
ARLINGTON, TX 76011-4511

03033817

PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS.

## 1 THIS IS AN APPLICATION FOR (Check appropriate item)

☐ A NEW LICENSE☐ B AMENDMENT TO LICENSE NUMBER☒ C RENEWAL OF LICENSE NUMBER

47-25327-01

## 2 NAME AND MAILING ADDRESS OF APPLICANT (Include ZIP code)

FibreK  
702 AFR Drive  
Fairmont, WV 26554

Docket No. 03033817

## 3 ADDRESS WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED

FibreK  
702 AFR Drive  
Fairmont WV 26554

## 4 NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION

Brad Morgan

## BUSINESS TELEPHONE NUMBER

304-333-6189

## BUSINESS CELLULAR TELEPHONE NUMBER

304-276-4001

## BUSINESS EMAIL ADDRESS

brad.morgan@resolutefp.com

SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE

## 5 RADIOACTIVE MATERIAL

a. Element and mass number; b. chemical and/or physical form, and c. maximum amount  
which will be possessed at any one time

## 6 PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED

Weigh Paper

## 8 TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS

7 INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR  
TRAINING EXPERIENCE

Brad Morgan 40 hour class

## 9 FACILITIES AND EQUIPMENT

Industrial Density Gauges

## 10 RADIATION SAFETY PROGRAM

## 11 WASTE MANAGEMENT

N/A

12 LICENSE FEES (Fees required only for new applications, with few exceptions\*)  
(See 10 CFR 170 and Section 170.31)

## FEE CATEGORY

2A

AMOUNT  
ENCLOSED \$

2,800.00

13 CERTIFICATION (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING  
UPON THE APPLICANT.

THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN  
CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, 36, 37, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT  
TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.  
WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 52 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO  
ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

## CERTIFYING OFFICER - TYPED/PRINTED NAME AND TITLE

Bradley E. Morgan

## SIGNATURE

Maintenance - Engineering MGR.

## DATE

4/24/15

## FOR NRC USE ONLY

DATE OF FEE	FEE LOG	FEE CATEGORY	AMOUNT RECEIVED	CHECK NUMBER	COMMENTS
-------------	---------	--------------	-----------------	--------------	----------

			\$		
--	--	--	----	--	--

APPROVED BY

DATE

## Suggested Format for Providing Information Requested in Items 5 Through 11 of NRC Form 313

**Table B.1 Items 5 & 6: Materials To Be Possessed and Proposed Uses**

Yes	No	Radioisotope	Manufacturer or Distributor Model No.	Quantity	Use As Listed on SSD Certificate	Specify Other Uses Not Listed on SSD Certificate
	✓	Cobalt-60	Sealed source manufacturer or distributor and model number: _____ Device manufacturer or distributor and model number: _____	Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate	Yes <input type="checkbox"/> Specific description of the gauge use: _____ _____ _____ _____ _____	<input type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: _____ (Submit safety analysis supporting safe use)
	✓	Krypton-85	Sealed source manufacturer or distributor and model number: _____ Device manufacturer or distributor and model number: _____	Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate	Yes <input type="checkbox"/> Specific description of the gauge use: _____ _____ _____ _____ _____	<input type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: _____ (Submit safety analysis supporting safe use)
	✓	Strontium-90	Sealed source manufacturer or distributor and model number: _____ Device manufacturer or distributor and model number: _____	Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate	Yes <input type="checkbox"/> Specific description of the gauge use: _____ _____ _____ _____ _____	<input type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: _____ (Submit safety analysis supporting safe use)
✓		Cesium-137	Sealed source manufacturer or distributor and model number: _____ Device manufacturer or distributor and model number: _____	Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate	Yes <input type="checkbox"/> Specific description of the gauge use: <u>see attachment</u> _____ _____ _____ _____ _____	<input type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: <u>Measure Paper</u> (Submit safety analysis supporting safe use)

APPENDIX B

Yes	No	Radioisotope	Manufacturer or Distributor Model No.	Quantity	Use As Listed on SSD Certificate	Specify Other Uses Not Listed on SSD Certificate
	✓	Americium-241	Sealed source manufacturer or distributor and model number: _____ Device manufacturer or distributor and model number: _____	Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate	Yes [ ] Specific description of the gauge use: _____ _____ _____ _____ _____	[ ] Not applicable [ ] Uses are: _____ (Submit safety analysis supporting safe use)
	✓	Other Isotope (Specify):	Sealed source manufacturer or distributor and model number: _____ Device manufacturer or distributor and model number: _____	Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate	Yes [ ] Specific description of the gauge use: _____ _____ _____ _____ _____	[ ] Not applicable [ ] Uses are: _____ (Submit safety analysis supporting safe use)
<i>Financial Assurance Required and Evidence of Financial Assurance Provided</i>						

**Table B.2 Items 7 Through 11: Training and Experience, Facilities and Equipment, Radiation Safety Program, and Waste Disposal**

Item No. and Title	Suggested Response	Yes	Alternative Procedures Attached
7. Individual(s) Responsible For Radiation Safety Program And Their Training And Experience	Before obtaining licensed materials, the proposed RSO will have successfully completed the training described in Criteria in the section entitled "Individual(s) Responsible for Radiation Safety Program and Their Training and Experience - Radiation Safety Officer" in NUREG-1556, Vol. 4, dated October 1998.	<input checked="" type="checkbox"/>	[ ]
7.1 Radiation Safety Officer Name: <u>Brad Morgan</u>	AND Before being named as the RSO, future RSOs will have successfully completed the training described in Criteria in the section entitled "Individual(s) Responsible for Radiation Safety Program and Their Training and Experience - Radiation Safety Officer" in NUREG-1556, Vol. 4, dated October 1998. Within 30 days of naming a new RSO, we will submit the new RSO's name to NRC to include in our license.		
7. Individual(s) Responsible For Radiation Safety Program And Their Training And Experience	PROPOSED AUTHORIZED USERS:  Before using licensed materials, authorized users will have successfully completed the training described in Criteria in the section entitled, "Authorized Users" in NUREG-1556, Vol. 4, dated October 1998.	<input checked="" type="checkbox"/>	[ ]
7.2 Authorized Users			
8. Training for Individuals Who in the Course of Employment are Likely to Receive Occupational Doses of Radiation in Excess of 1 mSv (100 mrem) in a Year (Occupationally Exposed Workers) and Ancillary Personnel	The applicant is <i>not</i> required to, and should not, submit is training program, for individuals who in the course of employment are likely to receive occupational doses of radiation in excess of 1 mSv (100 mrem) in a year (occupationally exposed workers) and ancillary personnel, to the NRC for review during the licensing phase.		Need Not Be Submitted with Application

APPENDIX B

Item No. and Title	Suggested Response	Yes	Alternative Procedures Attached
9. Facilities and Equipment	We will ensure that the location of each fixed gauge meets the Criteria in the section entitled "Facilities and Equipment" in NUREG-1556, Vol. 4, dated October 1998.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
10. Radiation Safety Program - Audit Program	The applicant is <i>not</i> required to, and should not, submit its audit program to the NRC for review during the licensing phase.	Need Not Be Submitted with Application	
10. Radiation Safety Program - Survey Instruments	<p>Surveys pursuant to 10 CFR 20.1501 will be performed by a person specifically authorized by the NRC or an Agreement State to perform these surveys.</p> <p style="text-align: center;"><b>OR</b></p> <p>We will use instruments that meet the Criteria in the section entitled "Radiation Safety Program - Instruments," in NUREG-1556, Vol. 4, dated August 1998, and <i>one</i> of the following:</p> <p>Each survey meter will be calibrated by the manufacturer or other person authorized by the NRC or an Agreement State to perform survey meter calibrations.</p> <p style="text-align: center;"><b>OR</b></p> <p>We will implement the model survey instrument calibration program in Appendix I to NUREG-1556, Vol. 4, dated October 1998.</p>	<input checked="" type="checkbox"/>	<input type="checkbox"/>
10. Radiation Safety Program - Material Receipt and Accountability	Physical inventories will be conducted at intervals not to exceed 6 months or at other intervals approved by the NRC, to account for all sealed sources and devices received and possessed under the license.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
10. Radiation Safety Program - Occupational Dosimetry	We will perform a prospective evaluation demonstrating that unmonitored individuals are not likely to receive, in one year, a radiation dose in excess of 10% of the allowable limits in 10 CFR Part 20 or we will provide dosimetry that meets the Criteria in the section entitled "Radiation Safety Program - Occupational Dosimetry," in NUREG-1556, Vol. 4, dated October 1998.	<input type="checkbox"/>	<input checked="" type="checkbox"/>

## APPENDIX B

Item No. and Title	Suggested Response	Yes	Alternative Procedures Attached
10. Radiation Safety Program - Public Dose	The applicant is not required to submit a response to the public dose section during the licensing phase. However, during NRC inspections, licensees must be able to provide documentation demonstrating, by measurement or calculation, that the total effective dose equivalent to the individual likely to receive the highest dose from the licensed operation does not exceed the annual limit for individual members of the public.	Need Not Be Submitted with Application	
10. Radiation Safety Program - Operating & Emergency Procedures	<p>If the gauge meets one or more of the safety conditions specified in "Discussion," in the section entitled "Radiation Safety Program-Operating Emergency Procedures," in NUREG 1556, Vol. 4, dated August 1998 state the following:</p> <p style="padding-left: 40px;">Operating and emergency procedures will be developed, implemented, maintained, and distributed, and will meet the Criteria in the section entitled "Radiation Safety Program - Operating and Emergency Procedures," in NUREG-1556, Vol. 4, dated August 1998.</p> <p>For each gauge requested that does not meet one or more of the safety conditions specified in "Discussion," in the section entitled "Radiation Safety Program-Operating Emergency Procedures," in NUREG 1556, Vol. 4, dated August 1998 provide your operating, emergency and lock-out (if applicable) procedures to NRC for review.</p>	<input checked="" type="checkbox"/>  <input type="checkbox"/> Procedures Attached	<input type="checkbox"/>
10. Radiation Safety Program - Leak Test	<p>Leak tests will be performed at intervals approved by the NRC or an Agreement State and specified in the Sealed Source and Device Registration Certificate. Leak tests will be performed by an organization authorized by NRC or an Agreement State to provide leak testing services for other licensees or using a leak test kit supplied by an organization authorized by NRC or an Agreement State to provide leak test kits to other licensees and according to the kit supplier's instructions.</p> <p style="text-align: center;"><b>OR</b></p> <p>We will implement the model leak test program published in Appendix M to NUREG-1556, Vol. 4, dated October 1998.</p>	<input checked="" type="checkbox"/>  <input type="checkbox"/>	<input type="checkbox"/>

APPENDIX B

Item No. and Title	Suggested Response	Yes	Alternative Procedures Attached
10. Radiation Safety Program - Maintenance	<p><u>ROUTINE MAINTENANCE</u> We will implement and maintain procedures for routine maintenance of our fixed gauges according to each manufacturer's or distributor's written recommendations and instructions.</p> <p><u>NON-ROUTINE MAINTENANCE OPERATIONS</u> The gauge manufacturer, distributor or other person authorized by NRC or an Agreement State will perform non-routine operations such as installation, initial radiation survey, repair, and maintenance of components related to the radiological safety of the gauge, gauge relocation, replacement, and disposal of sealed sources, alignment, or removal of a gauge from service.</p>	✓	<p>[ ]</p> <p>The information listed in Appendix N supporting a request to perform non-routing operations in-house is attached</p>
10. Radiation Safety Program - Transportation	The applicant is <i>not</i> required to submit its response to transportation during the licensing process; this issue will be reviewed during inspection. However, the licensee should develop, implement, and maintain transportation procedures according to NRC and DOT regulations.	Need Not Be Submitted with Application	
10. Radiation Safety Program - Fixed Gauges Used at Temporary Job Sites	<p>This is not applicable to our program. We will not use fixed gauges at temporary job sites.</p> <p style="text-align: center;"><b>OR</b></p> <p>We will develop, implement, maintain and distribute procedures that meet the Criteria in the section entitled "Radiation Safety Program - Fixed Gauges Used at Temporary Job Sites" in NUREG-1556, Vol. 4, dated October 1998.</p>	<p>✓ Not Applicable</p> <p>[ ]</p>	<p>[ ]</p>
10. Radiation Safety Program - Minimization of Contamination	The applicant is not required to submit a response to minimization of contamination if the applicant's responses meet the criteria for the following sections: Radioactive Material - Sealed Sources and Devices, Facilities and Equipment, Radiation Safety Program - Operating and Emergency Procedures, Radiation Safety Program - Leak Testing, and Waste Management - Gauge Transfer and Disposal.	Need Not Be Submitted with Application	

APPENDIX B

Item No. and Title	Suggested Response	Yes	Alternative Procedures Attached
11. Waste Management - Gauge Disposal & Transfer	The applicant is not required to submit a response to waste management during the licensing process. However, the licensee should develop, implement, and maintain gauge transfer and disposal procedures in its radiation protection program.		Need Not Be Submitted with Application



#### Compliance with dose limits for individual members of the general public

We currently possess 6 Ronan SA-1 source holders; 3 containing 50mCi CS-137 and 3 containing 5mCi CS-137. The source holders are mounted to frames over an elevated conveyor approximately 20 feet above the floor. Due to the elevation and field reduction from the conveyor the radiation levels below the conveyor are less than 0.2mR/hr. The radiation levels at 2 inches from the devices containing 50mCi are less than 2mR/hr; 0.3mR/hr at 12". The radiation level on the 5mCi devices at 2 inches is less than 0.3mR/hr.

Our employees are adjacent to the devices once every 6 months to perform shutter checks. This operation takes approximately 10 minutes per device. Working at 12 inches from the devices it is unlikely that any employee would receive a dose in excess of 2mrem/hr or 50mrem per year. Ref: 10CFR20.1302 or agreement state equivalent.

#### Question #5 List of Gauges

- 1) 0575GK SA8-F37 CS-137 50mCi
- 2) 0577GK SA1-F37 CS-137 50mCi
- 3) 0553GK SA1-F37 CS-137 50mCi
- 4) 5383CN SA1-F37 CS-137 5mCi
- 5) 5376CN SA8-F37 CS-137 5mCi
- 6) 5374CN SA1-F37 CS-137 5mCi

#### Question #6 Purpose – Weigh Paper Bales

Question #7 Responsible for Radiation Safety Program – Brad Morgan, 40 hour training class, refresher class in 2015

Question #8 Training – yearly training

Question #9 Facilities and equipment – Industrial density gauges

Question #10 Radiation Safety Program – We train employees yearly

Question #11 Waste Management – N/A

Question #12 License Fee Category 2A, Renewal Fee \$2,800