

Official Use Only-Security-Related Information

Close-out Survey Information

Stuart J. MacDonald – Radiation Safety Officer
SMAC Testing, Inc.
NRC License 21-32606-01-01

As we discussed by telephone, we need the following information before we can remove the "old" storage facility (311 Christy Way W, Saginaw, Michigan) from your license as a location of use.

1. Please provide a history of all the radionuclides used/stored at your "old" facility. For example, we used millicurie amounts of cesium-137 and americium-241/beryllium in sealed source form in Troxler Model 3440 moisture/density gauges.
2. Please submit a copy of the leak test results that was current for the sealed sources that were transferred from your "old" facility to your "new" facility. Also, please specify if you had any leaking sealed sources at your "old" use/storage facility. If you never had any leaking sealed sources, please state so.
3. Please confirm that all radionuclides have been removed and specify the date when all radionuclides were moved from your "old" facility and transferred to the "new" facility.

Please resubmit your request to remove the "old" use/storage facility and include the information above. Please include a cover letter on company letterhead, dated and signed (signed by an individual who is authorized to sign official documents on behalf of the licensee) with your response letter. Please call me at 630-829-9839 if you have any questions.

From the desk of:



Bill Reichhold

<p>The enclosed license document contains sensitive security-related information. When separated from this license document this Close-out Survey Information is uncontrolled</p>

Official Use Only-Security-Related Information