

U.S. NUCLEAR REGULATORY COMMISSION

TELEPHONE CONVERSATION RECORD

Date: 8/20/15

Time: 9:00 AM

Mail Control
or Report No(s).

588376

License No(s).

07-28657-01

Docket No(s).

030-32378

Name of Licensee: Mid-Delaware Imaging, Incorporated

Name of Participant(s): Justin Malecki, RSO
Janice Nguyen

Telephone No. 302-734-9888 Ext. 130

Subject: License Termination

(NOTE: This will be used as the Documents Title in ADAMS)

I reviewed the license termination package for Mid-Delaware Imaging, Incorporated, and all documentation is in order. I confirmed with Justin Malecki, the RSO, that the last patient treatment was in May 2015. All operations had ceased for at least one week prior to the closeout survey on June 4, 2015. The only byproduct material that was used was Tc-99m and the occasional use of I-123. I did ask that Mr. Malecki submit documentation of their survey meter calibration and their well counter efficiency.

Action Required: Wait for additional information and then terminate license.

Document Availability:

☒

Publicly Available

☐

Non-Publicly Available

☒

Non-Sensitive

☐

Non-Sensitive Copyright

☐

Sensitive- Proprietary

☐

Sensitive – Privacy Act (includes PII)

☐

Sensitive – Internal

☐

Sensitive – Security-Related

☐

Immediate Release

Normal Release Date:

Delayed Release Date:

SUNSI Review Completed By:

Janice Nguyen

/ RA /

Document Accession #: